



National Renewable Energy Laboratory

A national laboratory of the U.S. Department of Energy
Office of Energy Efficiency & Renewable Energy

Innovation for Our Energy Future

Evaluation of Radiometers in Full-Time Use at the National Renewable Energy Laboratory Solar Radiation Research Laboratory

S.M. Wilcox and D.R. Myers


Technical Report

NREL/TP-550-44627

December 2008

NREL is operated for DOE by the Alliance for Sustainable Energy, LLC

Contract No. DE-AC36-08-GO28308



Evaluation of Radiometers in Full-Time Use at the National Renewable Energy Laboratory Solar Radiation Research Laboratory

S.M. Wilcox and D.R. Myers


Prepared under Task Nos. CP09.3301 and PVD9.3110

Technical Report
NREL/TP-550-44627
December 2008

National Renewable Energy Laboratory
1617 Cole Boulevard, Golden, Colorado 80401-3393
303-275-3000 • www.nrel.gov

NREL is a national laboratory of the U.S. Department of Energy
Office of Energy Efficiency and Renewable Energy
Operated by the Alliance for Sustainable Energy, LLC

Contract No. DE-AC36-08-GO28308



NOTICE

This report was prepared as an account of work sponsored by an agency of the United States government. Neither the United States government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States government or any agency thereof.

Available electronically at <http://www.osti.gov/bridge>

Available for a processing fee to U.S. Department of Energy
and its contractors, in paper, from:

U.S. Department of Energy
Office of Scientific and Technical Information
P.O. Box 62
Oak Ridge, TN 37831-0062
phone: 865.576.8401
fax: 865.576.5728
email: <mailto:reports@adonis.osti.gov>

Available for sale to the public, in paper, from:

U.S. Department of Commerce
National Technical Information Service
5285 Port Royal Road
Springfield, VA 22161
phone: 800.553.6847
fax: 703.605.6900
email: orders@ntis.fedworld.gov
online ordering: <http://www.ntis.gov/ordering.htm>



Printed on paper containing at least 50% wastepaper, including 20% postconsumer waste

Preface

This report describes the evaluation of the relative performance of the complement of solar radiometers deployed at the National Renewable Energy Laboratory (NREL) Solar Radiation Research Laboratory (SRRL). The evaluation represents a limited sample of instruments (usually one) from various manufacturers, and thus should not be viewed as representative of a specific manufacturer or model of instrument, but only of the instruments at hand for the evaluation. We report an analysis of 12 global (total hemispherical) horizontal radiometers three direct beam pyrheliometers, and two diffuse pyranometers in full-time use at SRRL. Data for 12 months were retrieved from the SRRL Baseline Measurement System and included the test instrument measurements and data for a computed reference data set. Not all test instruments had a full 12-month period of record as noted in the report. This report offers no explicit conclusion about instrument performance, as it is aimed at a very general array of applications with a wide range of instrumentation and accuracy requirements.

Acknowledgments

The continually scrupulous work of Afshin Andreas, Beverly Kay, Peter Gotseff, and Ibrahim Reda in maintaining the complement of instruments at the Solar Radiation Research Laboratory and performing exacting and demanding outdoor radiometer calibrations is gratefully acknowledged. The leadership of Thomas Stoffel, Resource Integration Group Manager, and David Mooney, Center Director, in setting the high standards for operations of the Solar Radiation Research Laboratory is appreciated. The United States Department of Energy Solar Energy Technology Program supported this work under DOE prime contract number DE-AC36-99-GO10337.

Table of Contents

| | |
|--|-----------|
| List of Figures | vi |
| List of Tables | ix |
| 1.0 Introduction | 1 |
| 2.0 Approach | 3 |
| 2.1 Evaluation of Global Horizontal Pyranometers | 3 |
| 2.2 Evaluation of Direct Normal Pyrheliometers | 6 |
| 3.0 Results | 7 |
| 3.1 Pyranometer Sub-Hourly 1-Minute Data | 7 |
| 3.2 Averaged Data: Monthly and Annual Average Daily Totals | 7 |
| 3.3 Pyrheliometer and Cavity Pyrheliometer Comparisons | 13 |
| 4.0 Summary | 20 |
| 5.0 References | 22 |
| 6.0 Appendices | 23 |

List of Figures

| | |
|--|----|
| Figure 1-1. General SRRL configuration (left) with instrument platform (oriented north-south) located west of building. North is to the top, east to the right of the photograph. Instrument platform (right) with solar tracking instruments to the left (south) and horizontal instruments to right (north). | 1 |
| Figure 1-2. Schematic layout of relative positions of radiometers in the study, with representative photographs of the instruments as deployed. | 2 |
| Figure 3-1. Mean percent differences from reference global irradiance for pyranometers as function of zenith angle. See Appendix A, Table A1 for instrument list. | 9 |
| Figure 3-2. Percent standard deviations of differences from reference global irradiance for pyranometers as function of zenith angle. See Appendix A, Table A-1 for instrument list. | 10 |
| Figure 3-3. Summary of means and standard deviations of differences (in percent) between test radiometers and reference irradiance Monthly Mean Daily Total. See Appendix A, Table A-1 for instrument list..... | 10 |
| Figure 3-4. Percent difference between Monthly Mean Daily Totals of test radiometers and reference irradiance.. See Appendix A, Table A1 for instrument list..... | 12 |
| Figure 3-5. Difference between MMDT of Corrected and Uncorrected RSR and SPN1 Diffuse and reference diffuse. Note SPN1 data available only for months 8 to 11. | 13 |
| Figure 3-6. Pseudo time series plot of differences between reference AHF cavity radiometer and K&Z CH1 010256, Eppley NIP 31173E6, and Eppley NIP 257926E6. (Data in chronological order from individual calibration days which may be separated by periods from a week to many months; sometimes indicated by discontinuities.) | 15 |
| Figure 3-7. Differences in W/m^2 between reference AHF cavity radiometer and Kipp CH1 010256, Eppley NIP 31173E6, and Eppley NIP 25792E6 during broadband calibrations as function of time of day..... | 16 |
| Figure 3-8. Percent differences between the three test instrcuments and reference cavity pyrhelimeter as function of zenith angle bins during BORCAL. Means and standard deviations (1 sigma, also in percent) for each 10° zenith angle bin shown at center of bin. | 17 |
| Figure 3-9. Percent difference in Monthly Mean Daily Totals for NIP, SPN1, RSR Corrected and Uncorrected from the Kipp and Zonen CH1 010256..... | 18 |
| Figure 3-10. Average (dots) and standard deviation (1 sigma error bars) of percent difference between NIP 31137 and CH 1 by zenith angle over 11 months (169538 pts). Month 1 = July 2007; Month 11 = May 2008. | 19 |

| | |
|--|----|
| Figure 3-11. Histogram of percent differences between NIP 31137E6 and CH1 over all Zenith angles. Mean value of percent difference is 0.20%, standard deviation of 1.98% | 20 |
| Figure B-1. Uncorrected Eppley PSP: PSP 25782F3 | 26 |
| Figure B-1(a). One-minute data set percent differences; PSP Uncorrected | 26 |
| Figure B-1(b). Daily total percent difference; PSP Uncorrected..... | 26 |
| Figure B-2. Corrected Eppley PSP_C 25782F3 | 27 |
| Figure B-2(a). One-minute data set percent differences. PSP, Corrected. | 27 |
| Figure B-2(b). Daily total percent difference; PSP, Corrected. | 27 |
| Figure B-3. PSP Ventilated Uncorrected 28402F3..... | 28 |
| Figure B-3(a). One-minute data set percent differences; PSP, Ventilated, Uncorrected..... | 28 |
| Figure B-3(b). Daily total percent difference; PSP, Ventilated, Uncorrected. | 28 |
| Figure B-4. PSP Ventilated and Corrected PSP_V_C 28402F3..... | 29 |
| Figure B-4(a). One-minute data set percent differences; PSP, Ventilated, Corrected..... | 29 |
| Figure B-4(b). Daily total percent difference; PSP, Ventilated, Corrected. | 29 |
| Figure B-5. Kipp & Zonen CM22 (Ventilated) 010034 | 30 |
| Figure B-5(a). One-minute data set percent difference; CM22, Ventilated. | 30 |
| Figure B-5(b). Daily total percent difference; CM22, Ventilated. | 30 |
| Figure B-6. Kipp & Zonen CM3 010284..... | 31 |
| Figure B-6(a). One-minute data set percent differences; CM3..... | 31 |
| Figure B-6(b). Daily total percent difference; CM3. | 31 |
| Figure B-7. Kipp & Zonen CM6B 015189..... | 32 |
| Figure B-7(a). One-minute data set percent differences; CM6B..... | 32 |
| Figure B-7(b). Daily total percent difference; CM6B. | 32 |
| Figure B-8. Li-Cor LI 200 PY28257 | 33 |
| Figure B-8(a). One-minute data set percent differences; LI-200 PY. | 33 |
| Figure B-8(b). Daily total percent difference Li-200..... | 33 |
| Figure B-9. SP-LITE 970003..... | 34 |
| Figure B-9(a). One-minute data set percent differences; SP Lite Global. | 34 |
| Figure B-9(b). Daily total percent difference; SP Lite Global..... | 34 |
| Figure B-10. Irradiance, Inc. RSR2 –Global Uncorrected PY37627 | 35 |
| Figure B-10(a). One-minute data set percent differences; RSR2 Global Uncorrected..... | 35 |
| Figure B-10(b). Daily total percent difference; RSR2 Global Uncorrected. | 35 |
| Figure B-11. Irradiance, Inc. RSR2 global Corrected PY37627 | 36 |
| Figure B-11(a). One-minute data set percent differences; RSR2 Global Corrected..... | 36 |
| Figure B-11(b). Daily total percent difference RSR2 Global Corrected. | 36 |

| | |
|---|----|
| Figure B-12. TSP1 - 940703 | 37 |
| Figure B-12(a). One-minute data set percent differences; TSP1 Global. | 37 |
| Figure B-12(b). Daily total percent difference TSP1 Global..... | 37 |
| Figure B-13. TSP700 0212-02 (Ventilated)..... | 38 |
| Figure B-13(a). One-minute data set percent difference; TSP700 Global, Ventilated | 38 |
| Figure B-13(b). Daily total percent difference; TSP7000 Global, Ventilated. | 38 |
| Figure B-14. SPN1 GLOBAL A168 (4 months of data Feb 2008 -May 2008)..... | 39 |
| Figure B-14(a). One-minute data set percent differences; SPN1 Global. | 39 |
| Figure B-14(b). Daily total percent difference; SPN1 Global. | 39 |
| Figure B-15. SPN1 DIRECT A168 (4 months of data Feb 2008 -May 2008) | 40 |
| Figure B-15(a). One-minute data set percent differences; SPN1 Direct. | 40 |
| Figure B-15(b). Daily total percent difference; SPN1 Direct. | 40 |
| Figure B-16. RSR2 Direct Uncorrected (11 months of data Jul 2007 -May 2008)..... | 41 |
| Figure B-16(a). One-minute data set percent differences: RSR2 Direct Uncorrected. | 41 |
| Figure B-16(b). Daily total percent difference RS2 Direct Uncorrected..... | 41 |
| Figure B-17. RSR Direct Corrected (11 months of data Jul 2007 – May 2008) | 42 |
| Figure B-17(a). One-minute data set percent differences; RSR2 Direct Corrected. | 42 |
| Figure B-17(b). Daily total percent difference; RSR2 Direct Corrected. | 42 |
| Figure B-18. SPN1 Diffuse A168 (4 months of data Feb 2008 -May 2008)..... | 43 |
| Figure B-18(a). One-minute data set percent differences; SPN1 Diffuse. | 43 |
| Figure B-18(b). Daily total percent difference; SPN1 Diffuse. | 43 |
| Figure B-19. RSR2 Diffuse Corrected (11 months of data Jul 2007-May 2008) | 44 |
| Figure B-19(a). One-minute data set percent differences; RSR2 Diffuse Corrected. | 44 |
| Figure B-19(b). Daily total percent difference RSR2 Diffuse Corrected. | 44 |
| Figure B-20. RSR2 Diffuse Uncorrected (11 months of data Jul 2007 -May 2008)..... | 45 |
| Figure B-20(a). One-minute data set percent differences; RSR2 Diffuse Uncorrected. | 45 |
| Figure B-20(b). Daily total percent difference RSR2 Diffuse Uncorrected. | 45 |

List of Tables

| | |
|--|----|
| Table 3-1. Instrument performance by zenith angle (sub-hourly; one-minute), monthly mean daily total, and annual mean biases..... | 8 |
| Table 3-2. Table of statistics for percent differences in daily total irradiances versus reference irradiance..... | 11 |
| Table 3-3. Monthly Mean Daily Total Percent differences from Ref Global Horizontal | 12 |
| Table 3-4. Diffuse Horizontal RSR and SPN1 | 13 |
| Table 3-5. Kipp & Zonen 010256 CH1 Z angle statistics | 14 |
| Table 3-6. Eppler Nip 25792 Z angle statistics..... | 14 |
| Table 3-7. Eppler Nip 31137 Z angle statistics..... | 14 |
| Table 3-8. Direct Normal Comparison Eppler NIP, RSR and SPN1 with CH1 (mean percent difference in MMDT) by month; month 1 = JULY 2007; month 11 = MAY 2008..... | 18 |
| Table 3-9. Average percent differences between NIP 1 and CH 1 1-minute data by 10 degree zenith angle bin over 11 months (169538 pts) Month 1 = JULY 2007; Month 11 = MAY 2008 | 19 |
| Table A-1. The instruments evaluated and calibration uncertainties..... | 23 |
| Table B-1. Statistical summary 1minute data by zenith angle bin PSP 25782F3..... | 26 |
| Table B-2. Statistical summary 1-minute data by zenith angle bin PSP 25782F3 (Corrected).... | 27 |
| Table B-3. Statistical summary 1-minute data by zenith angle bin PSP 28402F3 (Ventilated, Uncorrected) | 28 |
| Table B-4. Statistical summary 1-minute data by zenith angle bin PSP 28402F3 (Ventilated, Corrected) | 29 |
| Table B-5. Statistical summary 1-minute data by zenith angle bin CM22 010034 (Ventilated).. | 30 |
| Table B-6. Statistical summary 1-minute data by zenith angle bin CM3 010284..... | 31 |
| Table B-7. Statistical summary 1-minute data by zenith angle bin CM6B 015189 | 32 |
| Table B-8. Statistical summary 1-minute data by zenith angle bin Li-200 PY28257 | 33 |
| Table B-9. Statistical summary 1-minute data by zenith angle bin SP-LITE 970003..... | 34 |

| | |
|--|-----------|
| Table B-10. Statistical summary 1-minute data by zenith angle bin RSR2 Global Uncorrected. | 35 |
| Table B-11. Statistical summary 1-minute data by zenith angle bin RSR2 Global Corrected..... | 36 |
| Table B-12. Statistical summary 1-minute data by zenith angle bin TSP1 Global | 37 |
| Table B-13. Statistical summary 1-minute data by zenith angle bin TSP700 Global | 38 |
| Table B-14. Statistical summary 1-minute data by zenith angle bin SPN1 Global..... | 39 |
| Table B-15. Statistical summary 1-minute data by zenith angle bin SPN1 Direct..... | 40 |
| Table B-16. Statistical summary 1-minute data by zenith angle bin RSR2 Direct Uncorrected.. | 41 |
| Table B-17. Statistical summary 1-minute data by zenith angle bin RSR2 Direct Corrected..... | 42 |
| Table B-18. Statistical summary 1-minute data by zenith angle bin SPN1 Diffuse..... | 43 |
| Table B-19. Statistical summary 1-minute data by zenith angle bin RSR2 Diffuse Corrected.... | 44 |
| Table B-20. Statistical summary 1-minute data by zenith angle bin RSR2 Diffuse Uncorrected. | 45 |

1 Introduction

The Solar Radiation Research Laboratory (SRRL) at the National Renewable Energy Laboratory (NREL) regularly collects measurements from approximately 100 solar and metrological instruments as part of its Baseline Measurement System (BMS). These measurements are available to the public within a few seconds of logging via the NREL Measurement and Instrumentation Data Center (http://www.nrel.gov/midc/srrl_bms). Figure 1-1 shows the general configuration of SRRL and the instrument platform (to left, or west, of building) where the instruments are deployed.



Figure 1-1. General SRRL configuration (left) with instrument platform (oriented north-south) located west of building. North is to the top, east to the right of the photograph. Instrument platform (right) with solar tracking instruments to the left (south) and horizontal instruments to right (north).

To support decision making for other projects, this report presents an analysis of 12 global (total hemispherical) horizontal radiometers, three direct beam pyrheliometers, and two diffuse pyranometers in full-time use at SRRL. Relative location of the radiometers is shown on the schematic layout of the SRRL instrument platform in Figure 1-2.

Data for 12 months were retrieved from SRRL's BMS and included the test instrument measurements and data for a reference data set computed from selected instruments. Not all test instruments had a full 12-month period of record as noted in the results. This report offers no explicit conclusion about instrument performance, as it is aimed at a very general array of applications with a wide range of instrumentation and accuracy requirements.

Note that the evaluation of these instruments represents a limited sample of instrument models (usually 1), and these results cannot be used with complete confidence to infer the same results or characteristics for all instruments of the same manufacturer and model. The evaluation used data from NREL's Solar Radiation Research Laboratory and results for instruments may be unique to the climate or geographical conditions at its Golden, Colorado, location (39.74 N latitude, 105.18 W longitude, and elevation 1829 m). Past experience indicates that many of these instruments respond to solar radiation to a different and usually unquantified degree in dissimilar physical environments.




Figure 1-2. Schematic layout of relative positions of radiometers in the study, with representative photographs of the instruments as deployed.

The evaluation is subdivided into two sections: pyrhemimeters (direct beam measurements) and pyranometers (global horizontal and diffuse horizontal measurements). The instruments evaluated are listed in Appendix A in Table A-1. The evaluation numerical results are summarized in Table 3-1. Detailed reports for each individual instrument are shown in Appendices B and C.

2 Approach

Two statistical summaries were prepared for each instrument, where sufficient data were available. Data for the 12 months from May 2006 to May 2007 were collected using the Solar Radiation Research Laboratory Baseline Measurement System. Data were recorded as 1-minute averages of 3 second (0.33 Hz, 20 samples per minute) sampled data using Campbell Scientific CR23X data loggers. This constituted the sub-hourly data set. Subsequently, the sub-hourly data sets were aggregated into hourly average data (mean of 60 1-minute averages), daily totals (sum of all sunup 1-minute data), and monthly mean daily totals (mean of the daily totals for all days in the month).

For the sub-hourly statistics, the values in the 80° to 90° bin are not included in the summary because of sensitivities to environmental error and general low irradiance conditions. The last three entries report the statistics for the Eppley Laboratories NIP 27592E6, 31137E6, and Kipp and Zonen CH1 010256 with respect to an absolute cavity radiometer [1], an Eppley Laboratories Automatic Hickey-Frieden (AHF) under clear sky conditions only. The monthly and annual statistics NA designation for the pyrhemimeters are due to the limited duration data set (see section 2.2 on pyrhemimeter evaluations).

2.1 Evaluation of Global Horizontal Pyranometers

Data for the portion of the study pertaining to pyranometers were collected under all sky conditions throughout a year (data quality filtering removed about 5% of the data, and some instruments as noted had a short period of record). Two independently evaluated instruments were used to create a computed global hemispherical irradiance data set against which the pyranometers' measured global hemispherical irradiance was compared. Two other independently evaluated instruments were used to validate and quality assess (QA) the reference data set and reject data inconsistent among those four instruments (computed global and QA instruments, respectively, see section 2.1.2). The computed global hemispherical (henceforth "Computed Global") and QA instruments are identified below.

2.1.1 Computed Global Irradiance

Computed from: Direct beam * COS(Z) + Diffuse

Direct Beam: Kipp & Zonen CH1 Serial No. 010256

Cal. Factor = 91.881 W/m²/mV

Responsivity (Rs) @45° = 10.884 μV/W/m²

CALIBRATION Uncertainty: +0.60% - 0.56% (for 30° ≤ Z ≤ 60°)

Diffuse: Eppley 8-48 Serial No. 32331

Cal Factor = 112.77 W/m²/mV

Responsivity (Rs) @45° = 8.868 μV/W/m²

CALIBRATION Uncertainty: +1.95% -1.89% (for 30° ≤ Z ≤ 60°)

2.1.2 Quality Assessment and Filtering of Reference Irradiance Data

The direct and diffuse irradiance measurements used to calculate the reference computed global irradiance data were evaluated against an independent global irradiance calculated from different direct beam and diffuse instruments listed below. SERI QC [2] was used to validate the measurements. SERI QC uses clearness index K_t for global, K_d for Diffuse, and K_n for direct beam (ratio of measured data to top of the atmosphere or extraterrestrial data of the same type, direct beam, and direct component on a horizontal) to check that the balance of the component sum equation $K_t = K_n + K_d$ is “true” within a tolerance of ± 0.03 units. SERI QC generated flags for each data set. Any 1-minute data record without a SERI QC flag of 1 (zenith angle $> 80^\circ$) or 3 (within the ± 0.03 tolerance limit) was rejected from the reference data set (approximately 5% of the data were rejected). In our experience, a flag of three indicates data from a well-maintained station.

QA Direct Beam = Eppley NIP Serial No. 31137E6

| Period | Serial # | Cal Fac W/m ² /mV | RS@45 μ V/W/m ² | Uncertainty ($30^\circ \leq Z \leq 60^\circ$) |
|-------------------------|----------|---------------------------------|-----------------------------------|--|
| 16-05-2006 – 08-06-2007 | 31137E6 | 117.78 | 8.4905 | +0.98/-0.92% |
| 08-06-2007 - 18-06-2008 | 31137E6 | 117.59 | 8.5043 | +1.11/-1.03% |
| 18-07-2008 – 31-05-2008 | 31137E6 | 117.61 | 8.5024 | +1.10/-1.00% |

QA Diffuse = Kipp & Zonen CM22 (Shading ball) radiometers:

| Period | Serial # | Cal Fac W/m ² /mV | RS@45 μ V/W/m ² | Uncertainty ($30^\circ \leq Z \leq 60^\circ$) |
|-------------------------|----------|---------------------------------|-----------------------------------|--|
| 01-06-2007 – 13-06-2007 | 010046 | 107.72 | 9.2832 | +1.24/-1.14% |
| 13-06-2007 – 23-05-2008 | 010034 | 91.704 | 10.905 | +1.81/-1.50% |
| 23-05-2008 – 31-05-2008 | 010046 | 107.56 | 9.2972 | +1.45/-1.67% |

The two diffuse instruments—8-48 and CM22—were chosen for the two roles of computing global irradiance and QA so as to remove the possibility of using the CM22 in computing reference irradiance for the global CM22.

2.1.3 Uncertainties of the Method

The evaluation data set was retrieved from the Solar Radiation Research Laboratory Baseline Measurement System recorded at 1-minute averages of 3 sec (0.33 Hz, 20 samples per minute) sampled data using Campbell Scientific CR23X data loggers:

01-06-2007 to 29-10-2007: CR23X s/n 3416 Factory Cal 01/25/2001
29-10-2007 to 31-05-2008: CR23X s/n 1162 Calibrated 01/19/2007

The manufacturer’s specified accuracy is $\pm 0.025\%$ of Full Scale Range (FSR typically ± 50 mV, meaning FSR is 100 mV) 0°C to 40°C . So in the 50 mV range: accuracy: ± 0.025 mV = 25 microvolts (μ V) with 1.67 μ V resolution. Thus, the data logger contribution to uncertainty is $\sim \pm 25$ μ V ± 1.67 μ V which is approximately ± 26.7 μ V.

Radiometer responsivities are generally $< \sim 5.0$ to $10.0 \mu\text{V}/\text{Wm}^{-2}$, i.e., 1.0 Wm^{-2} generates 5 to $10 \mu\text{V}$ of signal. The $26.7 \mu\text{V}$ logger uncertainty equates to approximately 5 watts or 2.7 watts (absolute) in BOTH reference and TEST instruments, or $\sim 1.0\%$ to 0.5% of reading at midrange values of 500 Wm^{-2} , and 2% or 1% of 1000 Wm^{-2} readings. We will be conservative and select the average of the larger uncertainties, 1.5% , for the data logger contribution to overall uncertainty.

Combined uncertainty in the computed reference irradiance is estimated from the weighted contribution of the beam and diffuse components, ranging from $\pm 1.0\%$ for extremely clear (nearly all direct beam) to $\pm 2.5\%$ overcast (all diffuse, or high zenith angles). The mean of the 250,981 1-minute reference uncertainties is $1.8\% \pm 0.5\%$. Total uncertainty in the comparisons is determined from the square root of the mean squared uncertainties, plus twice the standard deviation squared, plus the logger uncertainty contribution squared or $(1.8^2 + (2 \cdot 0.5)^2 + 1.5^2) = \pm 2.5\%$. Thus reference and data irradiances within $\pm 2.5\%$ of each other are considered identical.

For each 1-minute data point, global irradiance from each test instrument was compared with the concurrent computed global irradiance, and the difference between them calculated as a percent (test irradiance minus computed irradiance divided by computed irradiance). The resulting data set was used to calculate “sub-hourly” percent mean bias differences and standard deviations. The computed global irradiance data and test instrument global irradiance data were summed into daily total irradiance, and monthly means of daily total, which were compared by calculating percent differences as for the 1-minute data.

Table A-1 in Appendix A lists the instruments evaluated. In the case of the Kipp & Zonen CM22 global horizontal measurement, because of calibration and deployment schedules, two physical instruments were used for this evaluation. The table also shows the uncertainty in the responsivity (R_s) derived from NREL’s Broadband Outdoor Radiometer Calibration (BORCAL) Radiometer Calibration and Characterization (RCC) procedures [3]. Those uncertainties are assigned based on the range of responsivities computed during the BORCAL process between zenith angles of 30° and 60° . Since the R_s is specified at $Z = 45^\circ \pm 0.2^\circ$, the 30° to 60° Z angle range is used to indicate the variability in R_s over zenith angles with $\pm 15^\circ$ of the calibration point. Thus these uncertainties are a less comprehensive measure of performance than the results in this evaluation. The calibration results are represented as valid only for the environmental conditions existing on the day of the calibration, whereas this evaluation encompasses data and environmental conditions for an entire year. The calibration responsivity for each instrument was used to calculate irradiance for the instruments, but the calibration uncertainty of the test instruments was not used for this evaluation. It is provided to show the expected uncertainty in the determination of irradiances using the specified responsivity in the range of zenith angles from 30° to 60° .

2.1.4 Special Considerations

All data where the absolute value of the computed global irradiance was less than 10 Wm^{-2} , or differences between the computed irradiance and the test irradiance were $> 100\%$ were eliminated. This removed issues with computing the ratios of small irradiance values and transient effects from differing time constants which were seen to inordinately contribute to ‘percentage difference’ results. An additional 3% of the 250,000 data points were removed from the analysis using these criteria.

At the latitude of the SRRL, the minimum zenith angle in summer is 17 degrees, so there is no 0 to 10 degree zenith angle bin data.

The entire suite of instruments was maintained with daily cleaning five days per week by NREL staff. The effects of snow accumulation and frost in the wintertime may not have been completely removed from the data set. The QA process on the reference data set likely removed most snow effects for both the reference and test data set. But individual instruments may have accumulated snow at different rates under different conditions, and the effects of snow and frost may be present for some test instruments. Additionally, some instruments had the benefit of ventilators, which tends to mitigate frost and snow accumulation. These effects may show up as a negative or positive bias on certain days, increasing the scatter of data.

Zenith angle and thermal corrections for the Eppley PSP were developed by the National Renewable Energy Laboratory and are neither supplied by the manufacturer nor part of instrument output. The corrections require special calibrations and additional instrumentation (pyrgeometer) and were included in this study for consideration by the user. Corrections for the Irradiance Rotating Shadowband Radiometer (RSR2) are supplied by the manufacturer and corrected data are included as part of the instrument output.

2.2 Evaluation of Direct Normal Pyrheliometers

The absolute cavity radiometer (Eppley Automatic Hickey–Freidan) is the reference for NREL calibrations and for this work is presumed to be the pyrheliometer with the lowest uncertainty available [4]. Although a full year of data in varying conditions and seasons – such as that used for the pyranometer evaluations above – is desired, it is not practical to deploy the cavity full time because its aperture is open to the elements and internal components are subject to water or soiling.

For this evaluation we collected reference cavity radiometer data from all NREL calibration events from 2002 to 2007 and matched this reference data with three pyrheliometers from the SRRL baseline system (Eppley NIPs 25792E6 and 31137E6 and Kipp & Zonen CH1 010256). Because known and accepted uncertainties in the NREL calibrations are similar to expected differences among pyrheliometers, if there were different responsivities in use for an instrument in the different calibration events over the study period (e.g., see Table A-1), these responsivities were averaged. This removed noise effects of calibration uncertainty and also provided the opportunity to detect sensitivity drift with time (none were observed).

The portion of the study pertaining to pyrheliometers used only clear sky conditions (as mandated by the calibration events). Although the instruments are assumed to be linear in sensitivity with irradiance, overall performance could vary under different sky conditions. However, this may not detract from the usefulness of evaluation results presented here since performance under clear skies may be of greatest interest to solar power conversion applications.

2.2.1 Uncertainties for the Direct Normal Pyrheliometer Evaluation:

The NREL BORCAL process has a nominal uncertainty contribution from the cavity radiometer and low thermal and high precision data logging system (which is about five times more accurate than the routine monitoring data logger) of about 0.5%.

3 Results

3.1 Pyranometer Sub-Hourly 1-Minute Data

Table 3-1 and Figures 3-1 and 3-2 display a summary of the percent differences between the computed reference irradiances and the instrument data. The first two columns of Table 3-1 are for the sub-hourly (1-minute) data for each instrument or instrument configuration over all zenith angle ranges. The 3rd and 4th columns of Table 3-1 shows the percent bias error in monthly mean daily totals and annual average daily total data between the computed reference and the instrument data. Since the biases changed as a function of zenith angle (and hence as a function of the month of the year) a range of bias differences is shown, from minimum to maximum magnitude. Typically the minimum bias with respect to zenith angle is for the low zenith angle data (summer), and maximum bias occurs with the largest zenith angle bin (80 to 90 degrees) and winter data (highest average zenith angles for the year).

The majority of the pyranometers (9 of the 13) have a bias between $\pm 1\%$ to $\pm 5\%$ from the computed reference irradiance 1 minute data. These are the CM22 Ventilated, CM6B, PSP Ventilated and Corrected, TSP1, TSP700, SP LITE, LI200, and RSR2 Corrected and Uncorrected. Of these, four have a bias error less than equal to $\pm 2\%$ in the 1-minute data. These are the RSR2 Corrected, CM6B, CM22 Ventilated, and SP LITE. The standard deviation of the differences from the 1-minute reference irradiance are less than $\pm 6\%$ for the CM22 Ventilated, TSP700, RSR2 Corrected and Uncorrected, PSP Ventilated and Corrected, and LI200.

Two of the instruments (RSR2 and SPN1), while strictly pyranometer-based, provide direct normal irradiance (DNI) through calculation from the measured global and diffuse irradiances. DNI values from these instruments are included in the pyranometer results.

3.2 Averaged Data: Monthly and Annual Average Daily Totals

Of the 13 pyranometers, all but two have a mean bias difference from the computed reference irradiance Monthly Mean Daily Totals (MMDT) of less than $\pm 5\%$ all year long. The two exceptions are the PSP Uncorrected and PSP Ventilated Uncorrected, which underestimate winter irradiances by more than 7% due to poor cosine response with respect to the 45° zenith angle responsivity for these instruments.

In terms of annual average daily total (AADT) irradiance, all instruments agree with the computed reference irradiance annual average daily total irradiance within $\pm 3\%$, except the PSP Uncorrected, which has a mean bias of -3.9%. Eight of the instruments have less than $\pm 2\%$ bias from the computed reference irradiance AADT (CM22 Ventilated, SP Lite, TSP700, CM3, CM6B, LI200, RSR2 Corrected, and PSP Corrected.) Five of these have less than $\pm 1\%$ bias from the AADT of the computed reference irradiance (CM22 Ventilated, CM6B, LI200, RSR2 Corrected, and PSP Corrected).

Note there is insufficient coverage of the year by the Delta-T SPN1 radiometer to make an adequate comparison with the other pyranometers with respect to the annual daily totals. The MMDT for this instrument only span 4 months in the winter and spring seasons

Various combinations of mean bias error, standard deviation of the bias error, MMDT bias and AADT bias would result in different ‘rankings’ of the performance of the instruments with respect to the computed reference irradiance, which has an estimated uncertainty of $\pm 2.5\%$.

Figure 3-1 shows that for zenith angles less than 60° , *all of the pyranometer mean bias errors are less than $\pm 5\%$* . Figure 3-2 shows that for all but the CM3 pyranometer, the standard deviation of the differences from the computed reference irradiance is also *less than 5% for zenith angles less than 60°* .

Thus, it is imperative to note that the largest contribution to differences of all of the pyranometers with respect to the computed reference irradiance are due to issues with cosine response of the instruments for zenith angles greater than 60° .

Table 3-1. Instrument performance by zenith angle (sub-hourly; 1-minute), monthly mean daily total, and annual mean biases.

| Instrument (see Table A1 in Appendix A) | Sub-hourly Uncertainty % (range of bins to 80° Zenith) | | Mean Daily Total Uncertainty % | |
|---|--|----------------------|--|----------------------------|
| | Bias % max / min | Standard Deviation % | Monthly Bias % (monthly bins) max / min | Annual Bias ⁺ % |
| CM22_Vent | +0.6 to +0.4 | 2 to 3 | +0.77 to +0.18 | +0.48 |
| CM6B | +0.8 to -0.3 | 4 to 7 | +0.87 to -1.96 | +0.15 |
| CM3 | +0.1 to -7.4 | 2 to 7 | +0.37 to -1.84 | -1.23 |
| PSP_Vent_C | +0.0 to -3.1 | 2 to 6 | -0.06 to -5.13 | -2.08 |
| PSP_C ¹ | +5.7 to -0.3 | 2 to 11 | +1.17 to -3.27 | +0.79 |
| PSP_Vent_U | +1.6 to -11.3 | 2 to 8 | +0.01 to -8.38 | -3.88 |
| PSP_U | +2.5 to -7.5 | 2 to 9 | +0.42 to -10.3 | -2.78 |
| TSP700 | -0.8 to -2.5 | 4 to 6 | +1.50 to -0.01 | +1.09 |
| TSP1 | +3.2 to +0.4 | 2 to 12 | +4.59 to -0.41 | +2.32 |
| SPN1 (Glo)* | -0.3 to -3.7 | 4 to 7 | -0.67 to -6.04 | * |
| SPN1 (Dir)* | +8.1 to +3.0 | 19 to 24 | +15.4 to -7.90 | * |

¹ Infrared correction for unVentilated pyranometer using Ventilated pyrgeometer to measure net Infrared

| | | | | |
|-----------------|----------------|------------|----------------|-------|
| SPN1 (Dif)* | -13.8 to -4.3 | 7 to 11 | -12.9 to -10.4 | * |
| SP_LITE | +1.4 to +0.8 | 4 to 7 | +1.39 to +0.02 | +1.35 |
| LI200_TOT | +2.8 to -2.0 | 3 to 8 | +1.92 to -2.01 | +0.93 |
| RSR2_C (Glo)** | +1.0 to -1.2 | 4 to 6 | +0.53 to -2.54 | -0.73 |
| RSR2_C (Dir)** | -3.5 to -7.5 | 16 to 19 | -0.59 to -8.81 | -6.60 |
| RSR2_C (Dif) | -0.2 to +3.0 | 5 to 6 | -1.06 to +3.03 | +1.00 |
| RSR2_U (Glo)** | -0.8 to -2.5 | 4 to 6 | -0.81 to -3.59 | -2.17 |
| RSR2_U (Dir)** | +1.3 to +0.5 | 17 to 20 | -2.07 to +3.37 | +1.04 |
| RSR2_U (Dif)** | -12.0 to -15.0 | 13 to 16 | -19.3 to -19.6 | -14.1 |
| NIP 31137 DNI** | -0.2 to +0.4 | 1.7 to 2.1 | -0.30 to +1.20 | +0.10 |
| NIP 25792E6/Cav | +0.79 to -0.28 | 0.5 to 0.7 | N/A | N/A |
| NIP 31137E6/Cav | +0.46 to -0.53 | 0.4 to 0.6 | N/A | N/A |
| CH1 010256/Cav | +0.23 to +0.10 | 0.4 to 0.5 | N/A | N/A |

* Partial period of record (4 months); ** Partial period of record (11 months); + Average % difference in annual daily totals




Figure 3-1. Mean percent differences from reference global irradiance for pyranometers as function of zenith angle. See Appendix A, Table A-1 for instrument list.




Figure 3-2. Percent standard deviations of differences from reference global irradiance for pyranometers as function of zenith angle. See Appendix A, Table A-1 for instrument list.

Figure 3-3 is a bar chart of the mean difference between the MMDT of test and computed reference global (and the SPN1 Direct and Diffuse with respect to the CH1 direct and CM22 diffuse MMDT) irradiances. The error bars represent the standard deviation of the MMDT differences in percent of the mean irradiance.




Figure 3-3. Summary of means and standard deviations of differences (in percent) between test radiometers and reference irradiance Monthly Mean Daily Total. See Appendix A, Table A-1 for instrument list.

Table 3-2 shows the data of Figure 3-3 in tabular form, as well as the range of differences in MMDT in percent.

Table 3-2. Table of statistics for percent differences in daily total irradiances versus reference irradiance.

| | Count (N) | Mean | StDev (sd) | Maximum | Minimum | Range |
|----------------|--------------|---------------|---------------|---------|---------|-------|
| CM22_VENT_Pct | 366 | 0.480 | 0.788 | 5.8 | -4.6 | 10.4 |
| CM6B_Pct | 366 | 0.151 | 2.495 | 9.3 | -23.8 | 33.1 |
| CM3_Pct | 366 | -1.230 | 2.854 | 5.3 | -39.3 | 44.6 |
| PSP_VENT_C_Pct | 366 | -2.081 | 2.098 | 3.2 | -12.2 | 15.4 |
| PSP_C_Pct | 366 | 0.792 | 5.856 | 45.3 | -55.5 | 100.8 |
| PSP_VENT_U_Pct | 366 | -3.880 | 4.759 | 1.2 | -58.3 | 59.5 |
| PSP_U_Pct | 366 | -2.775 | 6.096 | 8.3 | -59.8 | 68.1 |
| TSP700_Pct | 366 | 1.086 | 2.177 | 7.1 | -35 | 42.1 |
| TSP1_Pct | 366 | 2.320 | 4.467 | 27.2 | -31.6 | 58.8 |
| GLO_SPN1_Pct | 152 | -2.928 | 3.121 | 6.8 | -14.7 | 21.5 |
| SP_LITE_Pct | 366 | 1.352 | 2.714 | 15.6 | -20.8 | 36.4 |
| LI200_Pct | 366 | 0.928 | 2.935 | 12.2 | -23.9 | 36.1 |
| DIR_SPN1_Pct* | 144 | 4.197 | 17.79 | 91.0 | -66.8 | 157.8 |
| RSR2_U (Glo)** | 341 | -2.174 | 2.812 | 11.5 | -35.7 | 47.2 |
| RSR2_U (Dir)** | 329 | 1.043 | 9.269 | 39.0 | -82.7 | 121.7 |
| RSR2_C (Glo)** | 341 | -0.733 | 3.015 | 14.2 | -47.4 | 47.7 |
| RSR2_C (Dir)** | 330 | -4.626 | 10.48 | 65.4 | -97.8 | 163.2 |

* Only 4 months of valid data (Feb 2008-May 2008); ** 11 months of data (July 07 to May 08)

Figure 3-3 and Table 3-2 summarize results for the aggregated months in the study. Figure 3-4 and Table 3-3 show the month to month variability of the differences in MMDT for the global hemispherical irradiances. Six units, the SPN1 Global, Corrected PSP, Corrected Ventilated PSP, Uncorrected PSP, Uncorrected Ventilated PSP, and TSP1 pyranometers exceed +/-2% difference for one or more months of the year, typically in the winter (month 7 is December, 2007) months. The TSP1 MMDT are 4% too high in the spring (Feb, Mar, Apr) months. The Corrected PSP MMDT difference is -3% only for one month (Dec). Note that the SPN1 Global data was present for January (month 8) to May (month 12).




Figure 3-4. Percent difference between Monthly Mean Daily Totals of test radiometers and reference irradiance.. See Appendix A, Table A1 for instrument list.

Table 3-3 shows the tabular data presented in Figure 3.4. Shaded cells highlight MMDT differences that exceed $\pm 2.5\%$.

Table 3-3. Monthly Mean Daily Total Percent differences from Ref Global Horizontal (CH1 + Diff)

| | JUN 2007 | JUL 2007 | AUG 2007 | SEP 2007 | OCT 2007 | NOV 2007 | DEC 2007 | JAN 2008 | FEB 2008 | MAR 2008 | APR 2008 | MAY 2008 |
|------------------------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|
| GLO_REF Wh/m ² | 6749 | 6473 | 5386 | 4503 | 3881 | 2306 | 2211 | 2500 | 3467 | 4597 | 6133 | 6001 |
| CM22_Vent | 0.266 | 0.203 | 0.184 | 0.319 | 0.372 | 0.205 | 0.064 | 0.317 | 0.527 | 0.771 | 0.612 | 0.687 |
| CM6B | 0.293 | 0.230 | 0.117 | 0.246 | 0.138 | 0.264 | -1.960 | -0.241 | 0.753 | 0.870 | 0.227 | 0.554 |
| CM3 | -1.606 | -1.244 | -1.585 | -1.342 | -1.207 | -1.843 | -1.638 | -0.581 | -0.195 | 0.386 | 0.355 | 0.233 |
| PSPVENC | -0.034 | -0.061 | -0.626 | -1.394 | -2.740 | -4.002 | -5.129 | -5.083 | -3.226 | -2.099 | -1.379 | -0.475 |
| PSP_C | 0.509 | 0.665 | 0.611 | 0.904 | 0.897 | -1.461 | -3.268 | 0.802 | 1.696 | 1.173 | 0.454 | 0.569 |
| PSPVENTU | -0.036 | 0.007 | -1.110 | -2.647 | -4.977 | -7.465 | -8.380 | -8.222 | -5.467 | -3.322 | -1.932 | -0.692 |
| PSPU | 0.214 | 0.423 | -0.419 | -1.446 | -3.355 | -8.293 | -10.314 | -5.995 | -3.011 | -1.354 | -0.833 | 0.180 |
| TSP700 | 0.670 | 0.305 | 0.544 | 0.755 | 0.831 | 1.253 | -0.010 | 1.201 | 1.488 | 1.500 | 0.858 | 0.571 |
| TSP1 | -0.209 | -0.313 | -0.270 | 0.423 | 1.337 | 2.719 | -0.411 | 3.793 | 4.589 | 4.597 | 3.657 | 2.169 |
| GLO_SPN1 | ---- | ---- | ---- | ---- | ---- | ---- | ---- | -6.040 | -4.043 | -1.705 | -0.668 | -0.979 |
| RSR_UNCOR | --- | -1.58 | -2.76 | -3.14 | -2.59 | -2.56 | -3.59 | -2.5 | -2.27 | -0.99 | -1.55 | -0.81 |
| RSR_Corr | --- | -1.13 | -2.42 | -2.54 | -1.42 | -0.48 | -0.78 | 0.08 | -0.32 | 0.53 | -0.30 | 0.34 |
| SP_LITE | 0.798 | 0.569 | 0.823 | 1.039 | 1.182 | 1.182 | 1.393 | 0.743 | 0.982 | 0.921 | 0.021 | 0.809 |
| LI200 | 1.802 | 1.639 | 1.308 | 0.686 | 0.524 | -0.350 | -2.011 | -0.208 | 0.893 | 1.628 | 1.701 | 1.918 |

As mentioned earlier, the largest contribution to the differences in the winter months is the deviation in response of the instruments as a function of the high zenith angles (greater than 60°) in the winter at the location of the Solar Radiation Research Laboratory.

The diffuse measurements provided by the RSR and SPN1 radiometers were evaluated separately for their relationship to the reference diffuse as shown in Table 3-4 and Figure 3-5.

Table 3-4. Diffuse Horizontal RSR and SPN1 (Mean % difference in monthly mean daily totals)*

| | JUN 2007 | JUL 2007 | AUG 2007 | SEP 2007 | OCT 2007 | NOV 2007 | DEC 2007 | JAN 2008 | FEB 2008 | MAR 2008 | APR 2008 | MAY 2008 |
|-----------------------------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|
| DIFFUSE REF Wh/m ⁻² | | 1992 | 1770 | 1365 | 1055 | 838 | 814 | 762 | 1214 | 1681 | 1842 | 2260 |
| RSR DIF U % | ---- | -13.22 | -13.42 | -16.10 | -19.35 | -18.82 | -16.19 | -17.88 | -13.14 | -10.85 | -11.61 | -6.35 |
| RSR DIF C % | ---- | -0.64 | -1.06 | -0.40 | 1.31 | 1.47 | 0.12 | 2.02 | 1.44 | 2.05 | 2.88 | 3.03 |
| SPN1 DIF % | ---- | ---- | ---- | ---- | ---- | ---- | ---- | ---- | -12.9 | -11.9 | -11.57 | -10.24 |

* Month 1 = July 2007; Month 11 = May 2008




Figure 3-5. Difference between MMDT of Corrected and Uncorrected RSR and SPN1 Diffuse and reference diffuse. Note SPN1 data available only for months 8 to 11.

The Corrected RSR Diffuse is typically within 3% of the reference Diffuse, however the Uncorrected RSR and SPN1 Diffuse readings are significantly low (more than -10%) for all months.

3.3 Pyrheliometer and Cavity Pyrheliometer Comparisons

3.3.1 Cavity pyrheliometer and test pyrheliometer comparison

The following plots and tables indicate the relative bias and scatter of the data among direct normal measuring instruments with respect to an absolute cavity radiometer reference during scheduled clear-sky calibration events. The uncertainty of the reference instrument (AHF Cavity radiometer) data is nominally $\pm 0.5\%$, meaning that the bias plots could move up or down by that magnitude and not be significant. Scatter in individual instrument plots within $\pm 0.5\%$ could be due as much to the reference instrument as the test instrument itself. However, movements in biases relative to other test instruments hold greater significance since each instrument was

compared against the same reference cavity radiometer data set. All zenith bins are 10 deg wide and the center of bin is the independent variable in plots. Units are percent (%) deviation from reference irradiance. Tables 3-5, 3-6, and 3-7 show statistics within each zenith angle bin for the differences between the test pyrheliometers and the reference cavity pyrheliometer.

Table 3-5. Kipp & Zonen 010256 CH1 Z angle statistics

| Zenith Angle Bin | Count (N) | Mean % | StDev % | Maximum % | Minimum % | Range % |
|---------------------------------|----------------------|-------------------|--------------------|----------------------|----------------------|--------------------|
| 15 | 879 | 0.234 | 0.396 | 1.241 | -0.97 | 2.22 |
| 25 | 2107 | 0.115 | 0.419 | 1.442 | -0.98 | 2.42 |
| 35 | 2107 | 0.115 | 0.419 | 1.442 | -0.98 | 2.42 |
| 45 | 2166 | 0.183 | 0.434 | 1.316 | -1.51 | 2.82 |
| 55 | 1847 | 0.150 | 0.441 | 1.342 | -2.30 | 3.64 |
| 65 | 1978 | 0.101 | 0.461 | 1.222 | -4.18 | 5.40 |
| 75 | 1815 | 0.185 | 0.452 | 1.348 | -3.12 | 4.47 |
| 85 | 1698 | 0.222 | 0.479 | 1.780 | -4.96 | 6.74 |

Table 3-6. Eppley Nip 25792 Z angle statistics

| Zenith Angle Bin | Count (N) | Mean % | StDev % | Maximum % | Minimum % | Range % |
|---------------------------------|----------------------|-------------------|--------------------|----------------------|----------------------|--------------------|
| 15 | 879 | -0.284 | 0.494 | 1.230 | -1.66 | 2.90 |
| 25 | 2107 | -0.168 | 0.478 | 1.197 | -1.81 | 3.00 |
| 35 | 2166 | -0.134 | 0.473 | 1.089 | -2.94 | 4.03 |
| 45 | 1847 | 0.0081 | 0.535 | 1.517 | -1.64 | 3.16 |
| 55 | 1978 | 0.1316 | 0.542 | 1.781 | -3.92 | 5.70 |
| 65 | 1815 | 0.3935 | 0.562 | 2.469 | -1.35 | 3.82 |
| 75 | 1697 | 0.7858 | 0.702 | 3.297 | -1.79 | 5.09 |
| 85 | 456 | 1.0388 | 0.799 | 2.828 | -1.21 | 4.04 |

Table 3-7. Eppley Nip 31137 Z angle statistics

| Zenith Angle Bin | Count (N) | Mean % | StDev % | Maximum % | Minimum % | Range % |
|---------------------------------|----------------------|-------------------|--------------------|----------------------|----------------------|--------------------|
| 15 | 879 | -0.528 | 0.351 | 0.422 | -2.46 | 2.88 |
| 25 | 2106 | -0.440 | 0.415 | 0.951 | -1.78 | 2.73 |
| 35 | 2166 | -0.218 | 0.473 | 1.400 | -1.93 | 3.33 |
| 45 | 1847 | -0.055 | 0.547 | 1.454 | -2.33 | 3.79 |
| 55 | 1978 | 0.058 | 0.554 | 1.936 | -4.69 | 6.62 |
| 65 | 1815 | 0.266 | 0.561 | 2.573 | -2.20 | 4.78 |
| 75 | 1698 | 0.458 | 0.603 | 2.665 | -4.42 | 7.09 |
| 85 | 456 | 0.475 | 0.821 | 2.976 | -1.48 | 4.46 |

Units of all comparisons are percent deviation from the absolute cavity (AHF) reference irradiance

The Direct Normal Irradiance (DNI) comparison was made with only values of the CH1 & NIP greater than 10 watts to remove very large percentage errors caused by ratios of small numbers.

Figures 3-6 and 3-7 show that the CH1 demonstrates flat zenith angle response with respect to the reference cavity radiometer and smaller standard deviation over the entire range of zenith angles. The two NIPs have significantly larger mean deviations, which vary by zenith angle; being minimal at low zenith angles and increasing at high zenith angles. The standard deviations of the two NIPs are comparable, and only slightly (0.1%) greater than the CH1 standard deviation.

Figure 3-6 shows the individual deviations from the reference absolute cavity pyrliometer. The normal baseline monitored data was merged with the cavity radiometer reference data during the clear sky calibration events, and percentage differences between the reference and test irradiances were computed.




Figure 3-6. Pseudo time series plot of differences between reference AHF cavity radiometer and K&Z CH1 010256, Eppler NIP 31173E6, and Eppler NIP 25792E6. (Data in chronological order from individual calibration days which may be separated by periods from a week to many months; sometimes indicated by discontinuities.)

Figure 3-7 is the same data as Figure 3-6, but plotted versus time of day for all of the days in Figure 3-6. Note the 'smile' in the NIP data, and rather flatter CH1 curve in Figure 3-7. Figure 3-

8 shows the variation of the mean percent difference (solid lines) between the reference cavity pyrliometer and the test pyrliometers (installed in the baseline monitoring system) as a function of zenith angle bin. The standard deviations of the differences as a function of the zenith angle bin are also shown with the dashed lines. The CH1 data are almost flat with respect to zenith angle, and the NIP data have increasing differences and standard deviations at increasing zenith angles. The apparent correlation with zenith angle may not be a first-order relationship. There is possibly some other factor causing varying differences in instrument output that also correlates with zenith angle.




Figure 3-7. Differences in W/m^2 between reference AHF cavity radiometer and Kipp CH1 010256, Eppeley NIP 31173E6, and Eppeley NIP 25792E6 during broadband calibrations as function of time of day.




Figure 3-8. Percent differences between the three test instruments and reference cavity pyrheliometer as function of zenith angle bins during BORCAL. Means and standard deviations (1 sigma, also in percent) for each 10° zenith angle bin shown at center of bin.

3.3.2 Test Direct Normal Data Comparison with CH1 Pyrheliometer

We conducted a separate study of the relationship between the Eppley, Kipp and Zonen, and three other direct beam irradiances derived from two other instruments over the period from June 2007 to May 2008. This study used the 1 minute monitored data. Eppley NIP 31137, Kipp and Zonen CH1 010256, direct beam from an Irradiance, Inc. rotating shadowband radiometer (RSR), and a Delta-T, Inc. SPN1 radiometer were studied. The RSR data is provided in two forms, with and without corrections for spectral and cosine effects. The correction scheme is provided by the manufacturer. The SPN1 direct beam data is computed from shaded (Diffuse) and unshaded (global) sensors mounted in the SPN1 instrument, using the manufacturer's calibration factor. The following table 3.8 and Figure 3.9 show percent differences from the CH1 data in monthly mean daily totals (MMDT) for the other DNI instruments

Table 3-8. Direct Normal Comparison Eppley NIP, RSR and SPN1 with CH1 (mean percent difference in MMDT) by month; month 1 = JULY 2007; month 11 = MAY 2008

| | JUN 2007 | JUL 2007 | AUG 2007 | SEP 2007 | OCT 2007 | NOV 2007 | DEC 2007 | JAN 2008 | FEB 2008 | MAR 2008 | APR 2008 | MAY 2008 |
|--|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|-------------|
| DIRECT CH1 MMDT Whm ⁻² | 7063 | 6109 | 5540 | 5322 | 5936 | 5021 | 4072 | 4922 | 4799 | 5140 | 6648 | 5189 |
| Eppley NIP % Δ | -0.33 | 1.15 | -0.28 | -0.18 | -0.03 | 0.42 | 0.05 | -0.21 | 0.27 | 0.39 | -0.33 | ---- |
| RSR DIR U % Δ | ---- | 3.37 | 2.74 | 1.46 | 1.47 | 0.45 | 0.76 | 2.99 | -2.07 | -1.46 | -0.64 | 1.80 |
| RSR DIR C % Δ | ---- | -0.59 | -3.74 | -4.58 | -4.19 | -5.13 | -3.86 | -3.90 | -7.28 | -8.81 | -4.60 | -4.00 |
| SPN1 DIR %Δ | ---- | ---- | ---- | ---- | ---- | ---- | ---- | ---- | -7.59 | +3.89 | +15.42 | +6.00 |




Figure 3-9. Percent difference in Monthly Mean Daily Totals for NIP, SPN1, RSR Corrected and Uncorrected from the Kipp and Zonen CH1 010256.

Table 3-9. Average percent differences between NIP 1 and CH 1 one-minute data by 10 degree zenith angle bin over 11 months (169538 pts) Month 1 = JULY 2007; Month 11 = MAY 2008

| Zenith angle bin | Count (N) | Mean % | StDev (sd) | Maximum % | Minimum % | Range |
|------------------|-----------|--------|------------|-----------|-----------|-------|
| 15 | 4447 | -0.20 | 1.71 | 9.93 | -3.91 | 13.84 |
| 25 | 14162 | 0.01 | 2.21 | 10.00 | -80.10 | 90.10 |
| 35 | 17701 | 0.11 | 2.13 | 10.00 | -26.89 | 36.89 |
| 45 | 21586 | 0.14 | 1.91 | 9.99 | -13.86 | 23.85 |
| 55 | 28766 | 0.23 | 1.96 | 10.00 | -62.44 | 72.44 |
| 65 | 36135 | 0.33 | 2.06 | 10.00 | -60.12 | 70.12 |
| 75 | 27069 | 0.37 | 1.83 | 10.00 | -14.41 | 24.41 |
| 85 | 19672 | 0.06 | 1.83 | 9.98 | -41.68 | 51.67 |




Figure 3-10. Average (dots) and standard deviation (1 sigma error bars) of percent difference between NIP 31137 and CH1 by zenith angle over 11 months (169538 pts).

Figure 3-11 shows a histogram of the percent difference between the NIP 31137 and the CH1 pyrheliometers for the 1-minute sub-hourly data over the 11 month period from July 2007 to May 2008. The mean difference is 0.2% with a standard deviation of $\pm 1.98\%$. This implies that over the year, the NIP and CH1 produce nearly identical average irradiances, but with a variation through the year of about $\pm 2\%$ (1 sigma) between the two. Figures 3.6 and 3.7 suggest the variation is mostly due to responsivity variations in the NIP instruments.




Figure 3-11. Histogram of percent differences between NIP 31137E6 and CH1 over all Zenith angles. Mean value of percent difference is 0.20%, standard deviation of 1.98%

4 Summary

This study shows that at the NREL SRRL location, for pyranometer data based on a single calibration factor (sometimes in combination with corrections for cosine response and infrared offsets) there is agreement to better than 5% for all of pyranometers for zenith angles in the range of 30° to 60° on a minute by minute basis. However, the various different cosine response curves for the radiometers, in conjunction with a single responsivity assignment (at 45° zenith angle $\pm 0.2^\circ$) leads to lack of agreement, exceeding 10% or 20% at 70° zenith angle, and more than 20% at 80° zenith angles. Correction and ventilation techniques applied to the data or instruments may not always improve the situation, but rather may add additional sources of uncertainty to the data. The $\pm 5\%$ range of agreement observed for zenith angles less than 60° is typical of the quoted uncertainty in sub-hourly pyranometer data over the past 30 years.

Averaging (hourly, daily totals and averages, monthly mean and annual mean daily totals) may help cancel out some of the random variability between instruments. However, as Figure 3.4 demonstrates, there is still the possibility of differences exceeding 5% among this family of instruments. There are differences (up to $\pm 2\%$) in the responsivity of pyrhemometers from the two different manufacturers' instruments evaluated here. An investigation is ongoing as to the sources of these differences. The design of the instruments and susceptibility to environmental influences are being studied. The $\pm 2\%$ differences observed among thermopile pyrhemometers,

including with respect to an absolute cavity pyrheliometer (basic uncertainty $\pm 0.5\%$) is in accordance with the long standing quote of $\pm 2\%$ uncertainty in direct normal irradiance instruments.

5 References

- [1] WMO (1996). *International Pyrheliometer Comparison VIII*. World Meteorological Organization, Geneva, Switzerland.
- [2], Maxwell, E.; Wilcox, S.; Rymes, M. (1993). Users Manual for SERI_QC Software: Assessing the Quality of Solar Radiation Data. NREL/TP-463-5608. Golden CO: National Renewable Energy Laboratory.
- [3] Myers, D. R.; Stoffel, T. L.; Reda, I.; Wilcox, S. M.; Andreas, A. M. (2002). "Recent Progress in Reducing the Uncertainty in and Improving Pyranometer Calibrations." *ASME Journal of Solar Energy Engineering* (124:1); pp. 44-50.
- [4] Reda, I.; Stoffel, T. L.; Myers, D. R. (1996). Calibration of a Solar Absolute Cavity Radiometer with Traceability to the World Radiometric Reference. NREL/TP-463-20619. Golden CO: National Renewable Energy Laboratory.

6 Appendices

Appendix A—Instrument List

Table A-1. The instruments evaluated and calibration uncertainties.

| Model | Manufacturer | Serial # | Installation Period | Rs @45 °V/W/m ² | CALIBRATION Uncertainty % [30° ≤ Z ≤ 60°] |
|--|-------------------|----------|----------------------|-------------------------------|--|
| CM22 (AC Ventilated) | Kipp & Zonen | 010034 | 06-01-07 to 06-13-07 | 10.896 | +1.85/-1.37% |
| | | 010046 | 06-13-07 to 05-23-08 | 9.3009 | +1.92/-1.17% |
| | | 010034 | 05-23-08 to 05-31-08 | 10.904 | +1.81/-1.50% |
| CM6B | Kipp & Zonen | 015189 | 06-01-07 to 06-08-07 | 10.871 | +2.53/-1.09% |
| | | | 06-08-07 to 05-31-08 | 10.912 | +1.98/-1.59% |
| CM3 | Kipp & Zonen | 010284 | 06-01-07 to 05-31-08 | 9.234 | ±5% (?) |
| PSP | Eppley Laboratory | 25782F3 | 06-01-07 to 06-13-07 | 8.8793 | +3.30/-4.49% |
| | | | 06-13-07 to 05-23-08 | 8.8568 | +2.98/-5.38% |
| | | | 05-23-08 to 05-31-08 | 8.7834 | +3.16/-5.67% |
| PSP Corrected¹ | Eppley Laboratory | 25782F3 | 06-01-07 to 06-13-07 | 9.0368 | +2.87/-3.92% |
| <i>Zenith correction and $V_c = V - W_{net} * RS_{net}$</i> | | | 06-13-07 to 05-23-08 | 9.0005 | +2.68/-4.71% |
| | | | 05-23-08 to 05-31-08 | 8.9422 | +2.78/-4.65% |
| PSP (AC Ventilated) | Eppley Laboratory | 28402F3 | 06-01-07 to 06-13-07 | 6.9504 | +2.24/-2.85% |
| | | | 06-13-07 to 05-23-08 | 6.9276 | +2.37/-2.88% |
| | | | 05-23-08 to 05-31-08 | 6.8796 | +2.09/-4.43% |
| PSP (AC Ventilated) Corrected | Eppley Laboratory | 28402F3 | 06-01-07 to 06-13-07 | 7.0901 | +1.87/-2.29% |
| <i>Zenith correction and $V_c = V - W_{net} * RS_{net}$</i> | | | 06-13-07 to 05-23-08 | 7.0557 | +2.05/-2.03% |
| | | | 05-23-08 to 05-31-08 | 7.0131 | +1.72/-3.56% |
| TSP700 (AC Ventilated) | YES, Inc. | 0212-02 | 06-01-07 to 06-13-07 | 2876.5 | +1.40/-1.04% |
| | | | 06-13-07 to 05-23-08 | 2908.5 | +1.22/-1.37% |
| | | | 05-23-08 to 05-31-08 | 2919.2 | +1.53/-1.66% |
| TSP1 | YES, Inc. | 940703 | 06-01-07 to 06-13-07 | 2004.1 | +1.91/-1.76% |
| | | | 06-13-07 to 05-23-08 | 2000.1 | +1.68/-2.44% |
| | | | 05-23-08 to 05-31-08 | 2026.7 | +1.93/-2.58% |
| SPN1 | Delta-T | A168 | 08-23-07 to 05-23-08 | 1000.0 | ±4% (?) |
| (GLO/DIR/DIF) | | | 05-23-08 to 05-31-08 | 976.01 | +4.30/-2.37% |

| Model | Manufacturer | Serial # | Installation Period | Rs @45 W/m ² | CALIBRATION Uncertainty % [30° ≤ Z ≤ 60°] |
|-------------------------|------------------|----------|-------------------------|----------------------------|--|
| SP_LITE | Kipp & Zonen | 970003 | 06-01-07 to 06-13-07 | 85.298 | +1.21/-1.96% |
| | | | 06-13-07 to 05-23-08 | 85.304 | +1.80/-1.75% |
| | | | 05-23-08 to 05-31-08 | 84.948 | +1.63/-2.05% |
| LI-200 | Li-Cor | PY28257 | 06-01-07 to 06-13-07 | 13.144 | +3.06/-4.50% |
| | | | 06-13-07 to 05-31-08 | 13.292 | +3.11/-5.51% |
| RSR2 Uncorrected | Irradiance, Inc. | PY37627 | Jun 07-May08 | 10.35 | +/- 5% (not stated) |
| RSR2 Corrected | (GLO/DIR/DIF) | PY37627 | Jun 07-May08 | 10.35 | +/-5% (not stated) |

¹Non-standard correction using ventilated pyrheliometer

Appendix B—Global and Diffuse Radiometer Evaluation Detailed Results

The following plots and tables indicate the relative bias and scatter of the data among instruments. The uncertainty of the reference instrument data is stated in this report as nominally $\pm 2.5\%$, meaning that the bias plots could move up or down by that magnitude and not be significant. Scatter in individual instrument plots within $\pm 2.5\%$ could be due as much to the reference instruments as the test instrument itself. However, movements in biases relative to other test instruments hold greater significant since each instrument was compared against the same reference data set.

The section below contains a statistical summary for each instrument, one instrument per page. For each instrument:

- Sub-hourly (1-minute) percent differences versus zenith angle, average and standard deviation percent differences in 10° wide zenith angle bins and percent differences in daily total irradiance between the reference (computed) global irradiance and test instrument data are shown as a function of zenith angle, zenith angle bin, and date, respectively. Error bars represent the standard deviations of the means. No data where the difference exceeded 100% in absolute value is included in the statistics mainly to avoid issues with small irradiance values. These time scale measurements are useful for many applications requiring analysis of the resource throughout the entire day.
- Table presenting the numerical data in the first graph.
- A presentation of the entire 1 minute data set of differences between measured and reference irradiances as a function of zenith angles (within plot limits of $\pm 25\%$ for graph clarity).
- Time series presentation of the daily total data as percent difference between test and reference instruments. This measurement is useful for applications requiring measurements on a daily time scale without regard to hourly or sub-hourly variations.




Figure B-1. Uncorrected Eppley PSP: PSP 25782F3

Table B-1. Statistical summary 1-minute data by zenith angle bin PSP 25782F3

| PSP_U | Count (N) | PSP Uncorrected % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|-------|-----------|-----------------------------|------------|---------|---------|-------|
| 15 | 5497 | 2.572 | 2.109 | 41.9 | -28.4 | 70.3 |
| 25 | 19111 | 1.729 | 4.995 | 32.4 | -72.6 | 105.0 |
| 35 | 24260 | 0.921 | 4.657 | 70.1 | -77.5 | 147.5 |
| 45 | 30018 | -0.232 | 4.410 | 29.9 | -77.1 | 107.0 |
| 55 | 39487 | -1.997 | 4.895 | 66.3 | -75.4 | 141.8 |
| 65 | 50754 | -5.047 | 8.973 | 69.3 | -79.9 | 149.2 |
| 75 | 40634 | -7.567 | 9.139 | 63.5 | -80.4 | 143.9 |
| 85 | 39182 | -17.357 | 22.251 | 99.8 | -99.9 | 199.7 |




Figure B-1(a). One-minute data set percent differences; PSP Uncorrected




Figure B-1(b). Daily total percent difference; PSP Uncorrected




Figure B-2. Corrected Eppley PSP (Corrected for Z angle response and IR offset) PSP_C 25782F3

Table B-2. Statistical summary 1-minute data by zenith angle bin PSP 25782F3 (Corrected)

| PSP_C | Count (N) | PSP Corrected (Z, IR) % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|-------|-----------|-----------------------------------|------------|---------|---------|-------|
| 15 | 5497 | 0.069 | 1.958 | 39.0 | -29.7 | 68.7 |
| 25 | 19111 | -0.284 | 4.822 | 30.4 | -73.1 | 103.5 |
| 35 | 24260 | -0.060 | 4.527 | 72.2 | -77.7 | 149.9 |
| 45 | 30018 | 0.372 | 4.443 | 32.6 | -77.2 | 109.8 |
| 55 | 39487 | 1.221 | 5.168 | 68.7 | -73.1 | 141.8 |
| 65 | 50753 | 0.877 | 9.774 | 79.2 | -76.5 | 155.7 |
| 75 | 40634 | 5.679 | 11.251 | 86.0 | -76.5 | 162.5 |
| 85 | 36205 | 26.580 | 25.118 | 100.0 | -96.1 | 196.0 |




Figure B-2(a). One-minute data set percent differences. PSP, Corrected.




Figure B-2(b). Daily total percent difference; PSP, Corrected.




Figure B-3. PSP Ventilated Uncorrected 28402F3

Table B-3. Statistical summary 1-minute data by zenith angle bin PSP 28402F3 (Ventilated, Uncorrected)

| PSP_V_U | Count (N) | PSP VENT UNCOR% Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|---------|-----------|---------------------------|------------|---------|---------|-------|
| 15 | 5497 | 1.603 | 2.503 | 39.1 | -28.1 | 67.2 |
| 25 | 19111 | 0.912 | 2.429 | 29.6 | -24.1 | 53.7 |
| 35 | 24260 | -0.148 | 2.711 | 64.3 | -28.0 | 92.3 |
| 45 | 30018 | -1.564 | 2.768 | 27.0 | -38.8 | 65.8 |
| 55 | 39487 | -3.390 | 3.229 | 62.5 | -53.5 | 116.0 |
| 65 | 50754 | -5.490 | 4.340 | 65.8 | -73.2 | 139.0 |
| 75 | 40614 | -11.305 | 7.629 | 19.0 | -99.6 | 118.5 |
| 85 | 33743 | -33.084 | 22.538 | 99.2 | -99.9 | 199.2 |




Figure B-3(a). One-minute data set percent differences; PSP, Ventilated, Uncorrected.




Figure B-3(b). Daily total percent difference; PSP, Ventilated, Uncorrected.




Figure B-4. PSP Ventilated and Corrected PSP_V_C 28402F3

Table B-4. Statistical summary 1-minute data by zenith angle bin PSP 28402F3 (Ventilated, Corrected)

| PSP_V_C | Count (N) | Sum | PSP VENT CORR Mean | StDev (sd) | Maximum | Minimum | Range |
|---------|-----------|----------|--------------------|------------|---------|---------|-------|
| 15 | 5497 | -621.3 | -0.113 | 2.039 | 36.9 | -29.0 | 65.9 |
| 25 | 19111 | -9083.12 | -0.475 | 1.943 | 28.5 | -25.3 | 53.9 |
| 35 | 24260 | -18038.3 | -0.744 | 2.077 | 67.3 | -27.0 | 94.3 |
| 45 | 30018 | -32552.2 | -1.084 | 2.191 | 30.2 | -34.7 | 65.0 |
| 55 | 39487 | -69002.7 | -1.747 | 2.679 | 64.9 | -44.3 | 109.2 |
| 65 | 50753 | -144207 | -2.841 | 3.416 | 67.4 | -55.0 | 122.4 |
| 75 | 40632 | -126642 | -3.117 | 6.090 | 33.6 | -99.5 | 133.1 |
| 85 | 38887 | -160636 | -4.131 | 17.940 | 100.0 | -99.9 | 199.9 |




Figure B-4(a). One-minute data set percent differences; PSP, Ventilated, Corrected.




Figure B-4(b). Daily total percent difference; PSP, Ventilated, Corrected.




Figure B-5. Kipp & Zonen CM22 (Ventilated) 010034

Table B-5. Statistical summary 1-minute data by zenith angle bin CM22 010034 (Ventilated)

| CM22 | Count (N) | Sum | CM22 % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|------|-----------|----------|------------------|------------|---------|---------|-------|
| 15 | 5497 | 2455.4 | 0.447 | 1.947 | 30.1 | -28.2 | 58.3 |
| 25 | 19111 | 11212.96 | 0.587 | 1.887 | 26.6 | -22.4 | 48.9 |
| 35 | 24260 | 15068.77 | 0.621 | 1.769 | 51.6 | -25.8 | 77.4 |
| 45 | 30018 | 15568.67 | 0.519 | 1.842 | 43.7 | -33.1 | 76.9 |
| 55 | 39487 | 20603.2 | 0.522 | 1.823 | 51.3 | -30.5 | 81.8 |
| 65 | 50754 | 18704.81 | 0.369 | 1.972 | 65.3 | -49.7 | 115.0 |
| 75 | 40634 | 25611.97 | 0.630 | 2.548 | 56.9 | -28.8 | 85.7 |
| 85 | 40339 | 188401.9 | 4.670 | 11.835 | 99.9 | -99.3 | 199.2 |




Figure B-5(a). One-minute data set percent difference; CM22, Ventilated.




Figure B-5(b). Daily total percent difference; CM22, Ventilated.




Figure B-6. Kipp & Zonen CM3 010284

Table B-6. Statistical summary 1-minute data by zenith angle bin CM3 010284

| CM3 | Count (N) | CM3 % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|-----|-----------|-----------------|------------|---------|---------|-------|
| 15 | 5495 | 0.101 | 4.799 | 77.5 | -58.5 | 136.0 |
| 25 | 19111 | 0.085 | 4.554 | 93.0 | -56.5 | 149.5 |
| 35 | 24260 | -0.140 | 4.523 | 71.1 | -53.9 | 125.0 |
| 45 | 30018 | -0.365 | 4.521 | 70.4 | -71.6 | 142.0 |
| 55 | 39487 | -0.967 | 4.315 | 54.8 | -56.1 | 110.8 |
| 65 | 50754 | -2.374 | 4.703 | 94.0 | -57.6 | 151.5 |
| 75 | 40633 | -7.415 | 7.491 | 59.0 | -70.4 | 129.4 |
| 85 | 34843 | -26.587 | 23.896 | 99.5 | -99.9 | 199.4 |




Figure B-6(a). One-minute data set percent differences; CM3.




Figure B-6(b). Daily total percent difference; CM3.




Figure B-7. Kipp & Zonen CM6B 015189

Table B-7. Statistical summary 1-minute data by zenith angle bin CM6B 015189

| CM6B | Count (N) | CM6B % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|------|-----------|------------------|------------|---------|---------|-------|
| 15 | 5497 | 0.852 | 2.346 | 40.6 | -28.1 | 68.7 |
| 25 | 19111 | 0.685 | 3.414 | 34.2 | -53.4 | 87.6 |
| 35 | 24260 | 0.574 | 4.209 | 78.2 | -70.0 | 148.3 |
| 45 | 30018 | 0.627 | 3.746 | 36.3 | -69.8 | 106.1 |
| 55 | 39487 | 0.515 | 3.736 | 72.1 | -62.0 | 134.1 |
| 65 | 50754 | -0.087 | 5.511 | 74.7 | -70.6 | 145.4 |
| 75 | 40634 | -0.338 | 6.831 | 63.2 | -71.9 | 135.1 |
| 85 | 40155 | -1.551 | 16.920 | 99.7 | -99.6 | 199.3 |




Figure B-7(a). One-minute data set percent differences; CM6B.




Figure B-7(b). Daily total percent difference; CM6B.




Figure B-8. Li-Cor LI 200 PY28257

Table B-8. Statistical summary 1-minute data by zenith angle bin Li-200 PY28257

| LI-200 | Count (N) | LI-200 % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|--------|-----------|--------------------|------------|---------|---------|-------|
| 15 | 5497 | 2.783 | 2.850 | 37.8 | -33.4 | 71.2 |
| 25 | 19111 | 2.799 | 3.086 | 28.4 | -30.5 | 58.9 |
| 35 | 24260 | 2.221 | 4.342 | 53.1 | -60.8 | 113.9 |
| 45 | 30018 | 1.474 | 5.017 | 29.7 | -64.0 | 93.7 |
| 55 | 39487 | 0.610 | 5.135 | 63.8 | -68.9 | 132.7 |
| 65 | 50754 | -0.801 | 6.601 | 61.6 | -74.3 | 135.9 |
| 75 | 40634 | -1.975 | 8.495 | 69.9 | -74.4 | 144.2 |
| 85 | 39972 | 2.906 | 17.091 | 99.9 | -98.8 | 198.8 |




Figure B-8(a). One-minute data set percent difference; LI-200 PY.




Figure B-8(b). Daily total percent difference Li-200.




Figure B-9. SP-LITE 970003

Table B-9. Statistical summary 1-minute data by zenith angle bin SP-LITE 970003

| SPLITE | Count (N) | SP-LITE % MeanDiff | StDev (sd) | Maximum | Minimum | Range |
|--------|-----------|--------------------|------------|---------|---------|-------|
| 15 | 5497 | 0.829 | 4.057 | 32.6 | -35.1 | 67.7 |
| 25 | 19111 | 1.214 | 4.019 | 32.3 | -31.7 | 64.0 |
| 35 | 24260 | 1.346 | 4.662 | 49.9 | -65.8 | 115.6 |
| 45 | 30018 | 1.429 | 5.146 | 34.6 | -66.5 | 101.2 |
| 55 | 39487 | 1.435 | 4.873 | 56.2 | -65.0 | 121.2 |
| 65 | 50754 | 1.147 | 5.864 | 68.5 | -78.8 | 147.3 |
| 75 | 40634 | 1.018 | 6.925 | 75.7 | -79.4 | 155.1 |
| 85 | 39685 | 12.479 | 19.067 | 100.0 | -58.3 | 158.2 |




Figure B-9(a). One-minute data set percent differences; SP Lite Global.

Figure B-9(b). Daily total percent difference; SP Lite Global.




Figure B-10. Irradiance, Inc. RSR2 –Global Uncorrected PY37627 (11 months Jul 2007 -May 2008)

Table B-10. Statistical summary 1-minute data by zenith angle bin RSR2 Global Uncorrected

| RSR2_U_G | Count (N) | RSR2 Uncorrected % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|----------|-----------|------------------------------|------------|---------|---------|-------|
| 15 | 3007 | -0.808 | 3.543 | 27.6 | -21.4 | 49.0 |
| 25 | 15555 | -1.135 | 3.526 | 28.2 | -60.0 | 88.3 |
| 35 | 21172 | -1.399 | 4.129 | 48.9 | -73.5 | 122.5 |
| 45 | 27137 | -1.703 | 4.125 | 27.6 | -79.9 | 107.5 |
| 55 | 37364 | -2.132 | 4.078 | 24.0 | -66.7 | 90.7 |
| 65 | 48789 | -2.539 | 4.831 | 38.6 | -51.4 | 89.9 |
| 75 | 38434 | -1.986 | 5.906 | 67.1 | -58.5 | 125.6 |
| 85 | 37203 | -1.090 | 12.613 | 99.6 | -96.4 | 196.0 |




Figure B-10(a). One-minute data set percent differences; RSR2 Global Uncorrected.




Figure B-10(b). Daily total percent difference; RSR2 Global Uncorrected.




Figure B-11. Irradiance, Inc. RSR2 global Corrected PY37627 (11 months Jul 2007 -May 2008)

Table B-11. Statistical summary 1-minute data by zenith angle bin RSR2 Global Corrected

| RSR2_C_G | Count (N) | RSR2 Corrected % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|----------|-----------|----------------------------|------------|---------|---------|-------|
| 15 | 3007 | 0.409 | 3.946 | 29.6 | -20.8 | 50.4 |
| 25 | 15555 | -0.312 | 3.863 | 30.8 | -60.0 | 90.8 |
| 35 | 21172 | -0.934 | 4.404 | 49.1 | -73.6 | 122.6 |
| 45 | 27137 | -1.221 | 4.350 | 28.2 | -79.9 | 108.1 |
| 55 | 37364 | -1.080 | 4.260 | 24.5 | -66.0 | 90.5 |
| 65 | 48789 | -0.351 | 5.008 | 41.6 | -49.6 | 91.2 |
| 75 | 38434 | 1.034 | 6.080 | 67.4 | -57.5 | 124.9 |
| 85 | 37287 | -4.035 | 19.442 | 99.4 | -98.7 | 198.1 |




Figure B-11(a). One-minute data set percent differences. RSR2 Global Corrected.

Figure B-11(b). Daily total percent difference RSR2 Global Corrected.




Figure B-12. TSP1 - 940703

Table B-12. Statistical summary 1-minute data by zenith angle bin TSP1 Global

| TSP1 | Count (N) | TSP1 % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|------|-----------|------------------|------------|---------|---------|-------|
| 15 | 5497 | 0.357 | 2.387 | 31.2 | -24.4 | 55.6 |
| 25 | 19111 | 0.952 | 3.949 | 30.7 | -67.3 | 98.0 |
| 35 | 24260 | 1.489 | 5.267 | 53.1 | -74.2 | 127.3 |
| 45 | 30018 | 1.757 | 4.898 | 36.5 | -74.0 | 110.4 |
| 55 | 39487 | 1.977 | 5.189 | 60.3 | -64.6 | 124.9 |
| 65 | 50754 | 1.693 | 8.685 | 73.8 | -100.0 | 173.8 |
| 75 | 40613 | 3.250 | 11.927 | 90.2 | -100.0 | 190.1 |
| 85 | 36888 | 8.469 | 29.463 | 100.0 | -100.0 | 200.0 |




Figure B-12(a). One-minute data set percent differences; TSP1 Global.




Figure B-12(b). Daily total percent difference TSP1 Global.




Figure B-13. TSP700 0212-02 (Ventilated)

Table B-13. Statistical summary 1-minute data by zenith angle bin TSP700 Global

| TSP700 | Count (N) | TSP 700 % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|--------|-----------|---------------------|------------|---------|---------|-------|
| 15 | 5497 | 0.058 | 3.990 | 29.9 | -51.9 | 81.8 |
| 25 | 19111 | 0.610 | 2.234 | 27.8 | -32.6 | 60.3 |
| 35 | 24260 | 0.750 | 2.146 | 49.4 | -44.1 | 93.4 |
| 45 | 30018 | 0.853 | 1.990 | 33.3 | -32.8 | 66.1 |
| 55 | 39487 | 0.867 | 2.756 | 52.9 | -73.9 | 126.8 |
| 65 | 50680 | 1.104 | 2.992 | 68.5 | -100.0 | 168.5 |
| 75 | 40564 | 2.358 | 3.432 | 69.8 | -95.1 | 164.9 |
| 85 | 39651 | 9.873 | 17.194 | 100.0 | -99.6 | 199.6 |




Figure B-13(a). One-minute data set percent difference; TSP700 Global, Ventilated

Figure B-13(b). Daily total percent difference; TSP7000 Global, Ventilated.




Figure B-14. SPN1 GLOBAL A168 (4 months of data Feb 2008 -May 2008)

Table B-14. Statistical summary 1-minute data by zenith angle bin SPN1 Global

| SPN1_G | Count (N) | MeanSPN1_Glo % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|--------|-----------|--------------------------|------------|---------|---------|-------|
| 15 | 862 | -0.308 | 4.983 | 15.4 | -15.7 | 31.1 |
| 25 | 7049 | -0.981 | 4.307 | 25.1 | -18.4 | 43.5 |
| 35 | 10244 | -1.297 | 4.489 | 23.7 | -91.9 | 115.6 |
| 45 | 13518 | -1.760 | 4.246 | 23.9 | -100.0 | 123.9 |
| 55 | 18292 | -2.733 | 4.354 | 54.4 | -100.0 | 154.4 |
| 65 | 20881 | -3.269 | 4.667 | 55.7 | -39.9 | 95.6 |
| 75 | 16840 | -3.656 | 7.362 | 67.4 | -35.8 | 103.3 |
| 85 | 15621 | 1.747 | 27.431 | 99.9 | -100.0 | 199.9 |




Figure B-14(a). One-minute data set percent differences; SPNI Global.




Figure B-14(b). Daily total percent difference; SPN1 Global.




Figure B-15. SPN1 DIRECT A168 (4 months of data Feb 2008 -May 2008)

Table B-15. Statistical summary 1-minute data by zenith angle bin SPN1 Direct

| SPN1_DIR | Count (N) | SPN1 DIRECT % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|----------|-----------|-------------------------|------------|---------|---------|--------|
| 15 | 672 | 4.143 | 22.4 | 99.25 | -99.99 | 199.24 |
| 25 | 5732 | 5.905 | 24.0 | 98.5 | -99.99 | 198.49 |
| 35 | 7996 | 8.057 | 21.1 | 99.82 | -99.99 | 199.81 |
| 45 | 10530 | 7.343 | 21.5 | 99.99 | -99.99 | 199.98 |
| 55 | 14304 | 5.489 | 20.1 | 99.81 | -99.99 | 199.8 |
| 65 | 16324 | 4.419 | 19.2 | 99.91 | -99.99 | 199.9 |
| 75 | 12410 | 3.024 | 20.4 | 99.71 | -99.99 | 199.7 |
| 85 | 9259 | -12.472 | 30.1 | 99.1 | -99.99 | 199.09 |




Figure B-15(a). One-minute data set percent differences; SPN1 Direct.




Figure B-15(b). Daily total percent difference; SPN1 Direct.




Figure B-16. RSR2 Direct Uncorrected (11 months of data Jul 2007 -May 2008)

Table B-16. Statistical summary 1-minute data by zenith angle bin RSR2 Direct Uncorrected

| RSR DIR U | Count (N) | RSR DIR UNCORRECTED % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|-----------|-----------|---------------------------------|------------|---------|---------|--------|
| 15 | 2599 | 1.251 | 16.8 | 99.09 | -98.08 | 197.17 |
| 25 | 17944 | 0.499 | 20.2 | 99.88 | -99.99 | 199.87 |
| 35 | 17944 | 0.499 | 20.2 | 99.88 | -99.99 | 199.87 |
| 45 | 22645 | 0.976 | 18.9 | 99.85 | -99.99 | 199.84 |
| 55 | 31499 | 1.022 | 17.6 | 99.64 | -99.99 | 199.63 |
| 65 | 40494 | 0.627 | 18.6 | 99.81 | -99.99 | 199.8 |
| 75 | 30449 | 2.875 | 20.4 | 99.74 | -99.99 | 199.73 |
| 85 | 22925 | -2.235 | 33.0 | 99.86 | -99.99 | 199.85 |




Figure B-16(a). One-minute data set percent differences: RSR2 Direct Uncorrected.




Figure B-16(b). Daily total percent difference RSR2 Direct Uncorrected.




Figure B-17. RSR Direct Corrected (11 months of data Jul 2007 – May 2008)

Table B-17. Statistical summary 1-minute data by zenith angle bin RSR2 Direct Corrected

| RSR DIR C | Count (N) | RSR DIR CORRECTED % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|-----------|-----------|-------------------------------|------------|---------|---------|--------|
| 15 | 2504 | -3.507 | 17.0 | 97.41 | -99.51 | 196.92 |
| 25 | 12387 | -6.899 | 19.4 | 95.29 | -99.99 | 195.28 |
| 35 | 16643 | -7.509 | 18.4 | 92.79 | -99.99 | 192.78 |
| 45 | 20957 | -7.255 | 17.1 | 99.67 | -99.97 | 199.64 |
| 55 | 29369 | -6.606 | 15.9 | 94.73 | -99.99 | 194.72 |
| 65 | 37428 | -6.189 | 16.3 | 94.61 | -99.97 | 194.58 |
| 75 | 27961 | -5.025 | 16.5 | 86.66 | -99.98 | 186.64 |
| 85 | 18818 | -9.558 | 25.1 | 99.62 | -99.99 | 199.61 |




Figure B-17(a). One-minute data set percent differences; RSR2 Direct Corrected.




Figure B-17(b). Daily total percent difference; RSR2 Direct Corrected.




Figure B-18. SPN1 Diffuse A168 (4 months of data Feb 2008 -May 2008)

Table B-18. Statistical summary 1-minute data by zenith angle bin SPN1 Diffuse

| SPN1 DIF | Count (N) | SPN1 Diffuse % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|----------|-----------|--------------------------|------------|---------|---------|--------|
| 15 | 862 | -4.277 | 6.5 | 20.43 | -30.08 | 50.51 |
| 25 | 7049 | -6.911 | 7.0 | 23.23 | -37.16 | 60.39 |
| 35 | 10244 | -10.039 | 7.7 | 16.03 | -88.52 | 104.55 |
| 45 | 13517 | -10.966 | 8.0 | 22.6 | -47.6 | 70.2 |
| 55 | 18291 | -11.910 | 8.2 | 44.86 | -39.34 | 84.2 |
| 65 | 20881 | -13.658 | 9.1 | 48.6 | -57.41 | 106.01 |
| 75 | 16840 | -13.873 | 10.9 | 27.11 | -49.79 | 76.9 |
| 85 | 15959 | -0.635 | 22.3 | 99.96 | -96.68 | 196.64 |




Figure B-18(a). One-minute data set percent differences; SPN1 Diffuse.

Figure B-18(b). Daily total percent difference; SPN1 Diffuse.




Figure B-19. RSR2 Diffuse Corrected (11 months of data Jul 2007-May 2008)

Table B-19. Statistical summary 1-minute data by zenith angle bin RSR2 Diffuse Corrected

| RSR DIFF C | Count (N) | RSR DIFF CORRECTED % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|------------|-----------|--------------------------------|------------|---------|---------|--------|
| 15 | 2997 | 0.37 | 5.95 | 39.63 | -78.50 | 118.13 |
| 25 | 15534 | 0.46 | 5.68 | 99.89 | -90.08 | 189.97 |
| 35 | 21144 | -0.23 | 6.10 | 93.37 | -98.45 | 191.82 |
| 45 | 27098 | -0.19 | 5.40 | 74.27 | -99.11 | 173.38 |
| 55 | 37353 | 0.33 | 4.77 | 61.93 | -92.93 | 154.86 |
| 65 | 48783 | 1.09 | 5.23 | 99.94 | -96.01 | 195.95 |
| 75 | 38433 | 3.23 | 6.32 | 61.74 | -68.33 | 130.07 |
| 85 | 37406 | -1.06 | 19.47 | 99.47 | -100.00 | 199.47 |




Figure B-19(a). One-minute data set percent differences; RSR2 Diffuse Corrected.




Figure B-19(b). Daily total percent difference RSR2 Diffuse Corrected.




Figure B-20. RSR2 Diffuse Uncorrected (11 months of data Jul 2007 -May 2008)

Table B-20. Statistical summary 1-minute data by zenith angle bin RSR2 Diffuse Uncorrected.

| RSR DIF U | Count (N) | RSR DIFF UNCORRECTED % Mean Diff | StDev (sd) | Maximum | Minimum | Range |
|-----------|-----------|----------------------------------|------------|---------|---------|--------|
| 15 | 2996 | -14.69 | 13.96 | 20.71 | -92.50 | 113.21 |
| 25 | 15526 | -12.96 | 13.61 | 98.08 | -92.68 | 190.76 |
| 35 | 21129 | -14.19 | 14.38 | 81.66 | -99.58 | 181.24 |
| 45 | 27087 | -15.31 | 15.12 | 91.72 | -99.29 | 191.01 |
| 55 | 37349 | -17.75 | 15.82 | 32.48 | -99.10 | 131.58 |
| 65 | 48783 | -17.89 | 15.94 | 98.54 | -82.18 | 180.72 |
| 75 | 38432 | -15.34 | 14.57 | 56.72 | -94.77 | 151.49 |
| 85 | 37227 | -5.90 | 15.48 | 99.72 | -99.64 | 199.36 |




Figure B-20(a). One-minute data set percent differences; RSR2 Diffuse Uncorrected.




Figure B-20(b). Daily total percent difference RSR2 Diffuse Uncorrected.

REPORT DOCUMENTATION PAGE

Form Approved
OMB No. 0704-0188

The public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing the burden, to Department of Defense, Executive Services and Communications Directorate (0704-0188). Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.

PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ORGANIZATION.

| | | | | | | |
|---|-----------------------------|------------------------------|-------------------------------------|--|---|--|
| 1. REPORT DATE (DD-MM-YYYY) December 2008 | | | 2. REPORT TYPE Technical Report | | 3. DATES COVERED (From - To) | |
| 4. TITLE AND SUBTITLE Evaluation of Radiometers in Full-Time Use at the National Renewable Energy Laboratory Solar Radiation Research Laboratory | | | | 5a. CONTRACT NUMBER DE-AC36-08-GO28308 | | |
| | | | | 5b. GRANT NUMBER | | |
| | | | | 5c. PROGRAM ELEMENT NUMBER | | |
| 6. AUTHOR(S) S.M. Wilcox and D.R. Myers | | | | 5d. PROJECT NUMBER NREL/TP-550-44627 | | |
| | | | | 5e. TASK NUMBER PVD9.3110; CP09.3310 | | |
| | | | | 5f. WORK UNIT NUMBER | | |
| 7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) National Renewable Energy Laboratory 1617 Cole Blvd. Golden, CO 80401-3393 | | | | 8. PERFORMING ORGANIZATION REPORT NUMBER NREL/TP-550-44627 | | |
| 9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) | | | | 10. SPONSOR/MONITOR'S ACRONYM(S) NREL | | |
| | | | | 11. SPONSORING/MONITORING AGENCY REPORT NUMBER | | |
| 12. DISTRIBUTION AVAILABILITY STATEMENT National Technical Information Service U.S. Department of Commerce 5285 Port Royal Road Springfield, VA 22161 | | | | | | |
| 13. SUPPLEMENTARY NOTES | | | | | | |
| 14. ABSTRACT (Maximum 200 Words) This report describes the evaluation of the relative performance of the complement of solar radiometers deployed at the National Renewable Energy Laboratory (NREL) Solar Radiation Research Laboratory (SRRL). This report presents an analysis of 12 global (total hemispherical) horizontal radiometers, three direct beam pyrheliometers, and two diffuse pyranometers in full-time use at SRRL. Data for 12 months were retrieved from the SRRL Baseline Measurement System and included the test instrument measurements and data for a computed reference data set. This report offers no explicit conclusion about instrument performance, as it is aimed at a very general array of applications with a wide range of instrumentation and accuracy requirements. | | | | | | |
| 15. SUBJECT TERMS SRRL, Solar Radiation Research Laboratory, radiometers, pyrheliometers, pyranometers; Baseline Measurement System; BSM | | | | | | |
| 16. SECURITY CLASSIFICATION OF: | | | 17. LIMITATION OF ABSTRACT UL | 18. NUMBER OF PAGES | 19a. NAME OF RESPONSIBLE PERSON | |
| a. REPORT Unclassified | b. ABSTRACT Unclassified | c. THIS PAGE Unclassified | | | 19b. TELEPHONE NUMBER (Include area code) | |

Standard Form 298 (Rev. 8/98)
Prescribed by ANSI Std. Z39.18