

Traceable Pyrgometer Calibrations

NREL: Mike Dooraghi, Mark Kutchenreiter, Ibrahim Reda, Aron Habte, Manajit Sengupta, Afshin Andreas, Martina Newman
SGP RCF: Craig Webb

Abstract: The Atmospheric Radiation Measurement (ARM) program provides high-quality radiometric data from approximately 150 instruments deployed at Southern Great Plains (SGP), Eastern North Atlantic (ENA), North Slope of Alaska (NSA), Oilkrok (OL), and the ARM Mobile Facilities (AMF) sites. These instruments are deployed on the Solar Infrared Radiation Station (SIRS), SKYRAD, and GNRAD instrument platforms. In addition to the operational radiometers, there are more than 200 other radiometers that are calibrated and used for instrument swaps and replacements. The National Renewable Energy Laboratory (NREL) and ARM, through the Radiometric Calibration Facility (RCF) at SGP, provides Broadband Outdoor Radiometer Calibrations (BORCAL) for all shortwave (SW) and longwave (LW) radiometers that are deployed by the ARM program. The BORCAL-SW is traceable to the International System of Units (SI) through the World Radiometric Reference (WRR). On the other hand, the SI standard is not yet established for longwave measurements. Both NREL and ARM continue to improve radiometric measurements through the introduction of new methods to reduce uncertainty in calibration and field measurements. A significant part of this effort is to establish the longwave traceability to SI units and deploy the BORCAL-LW using an interim World Infrared Standard Group (WISG) for traceability. Deploying the BORCAL-LW capability has been performed under ARM program ECO-00781, "Establish Pyrgometer Calibrations Traceable to the WISG." The stated purpose of the ECO is to adopt the consensus WISG for calibrating pyrgometers used by ARM for broadband longwave irradiance data collected from SIRS, SKYRAD, and GNRAD instrument platforms. This poster presents the development, implementation, and operation of the BORCAL-LW system at the SGP RCF for the calibration of pyrgometers that provide traceability to WISG.

Calibration Certificate from BORCAL 2016-02 for PIR 30133F3




Figure 1. Page 1 of the calibration certificate contains the customer and pyrgometer information. It also contains data acquisition and reference pyrgometer traceability information along with the calibration procedure and the technical manager's signature.

Figure 2. Page 2 of the calibration certificate contains the measurement equation with calibration coefficients and constants necessary to calculate the irradiance of the PIR to the sky. Page 2 also contains the residual graphs (as compared to the reference irradiance), calibration coefficients, and uncertainty results for both $k=0$ and $k < 0$.




Figure 5. The residual in W/m^2 compared to the reference instruments ranges from -1.8 to 7.2 for the control instrument (36368F3) and from -1.3 to 1.3 for the measurement assurance standard (30133F3) when using the manufacturer's calibration. The residual in W/m^2 compared to the reference instruments ranges from -1.3 to 1.3 for the control instrument (36368F3) and from -1.6 to 1.6 for the measurement assurance standard (30133F3) when using the BORCAL-LW calibration.

Control Instrument History




Figure 7. Results from the eight BORCAL-LW events for the control instrument (36368F3) and the measurement assurance standard (30133F3) are shown above.



Figure 6. Craig Webb installing pyrgometers on one of the LW trackers at the SGP RCF.

Metrics for LW calibrations from July 2015 to March 2016

- Total number of instruments: **92**
- Total individual instruments: **81**
- Total failed calibrations: **8** (Most of these were later linked to a failed channel.)
- Average uncertainty: $\pm 2.31 W/m^2$
- Highest uncertainty: $\pm 2.59 W/m^2$
- Lowest uncertainty: $\pm 2.26 W/m^2$
- Average number of successful calibrations per session: **11.75**
- Average length of time for BORCAL-LW to complete: **30.6 days**
- Shortest BORCAL-LW event: **17 days**
- Longest BORCAL-LW event: **44 days**
- Current range of K0 values for all successfully calibrated instruments: **-9.6467 to 4.0452**
- Current range of K1 values for all successfully calibrated instruments: **0.2152 to 0.34601**
- Current range of K2 values for all successfully calibrated instruments: **0.9929 to 1.0253**
- Current range of K3 values for all successfully calibrated instruments: **-5.1028 to 0.00087**




Figure 4. The BORCAL-LW can be concluded once data has been collected under all sky conditions. This allows for calibration throughout the full voltage output range. The above graph shows the residuals in between the reference irradiance and the instrument irradiance using the factory calibration.

Documentation of the BORCAL-SW and BORCAL-LW processes conducted at SGP can be found here: <http://www.nrel.gov/docs/fy15osti/65035.pdf>
Documentation of the Radiometer Calibration and Characterization (RCC) software can be found here: <http://www.nrel.gov/docs/fy16osti/65844.pdf>

This work was supported by Argonne National Laboratory MPO No. 2T-30084 and the U.S. Department of Energy under Contract No. DE-AC36-08-GO28308 with the National Renewable Energy Laboratory.