



NREL Comparison of Absolute Cavity Pyrometers and Pyrometers Traceable to the World Infrared Standard Group and the Infrared Integrating Sphere: September 23-October 4, 2019

Ibrahim Reda, Afshin Andreas, Peter Gotseff,
Mark Kutchenreiter, and Martina Stoddard

National Renewable Energy Laboratory

**NREL is a national laboratory of the U.S. Department of Energy
Office of Energy Efficiency & Renewable Energy
Operated by the Alliance for Sustainable Energy, LLC**

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at www.nrel.gov/publications.

Contract No. DE-AC36-08GO28308

Technical Report
NREL/TP-1900-75119
October 2019



NREL Comparison of Absolute Cavity Pyrometers and Pyrometers Traceable to the World Infrared Standard Group and the Infrared Integrating Sphere: September 23-October 4, 2019

Ibrahim Reda, Afshin Andreas, Peter Gotseff,
Mark Kutchenreiter, and Martina Stoddard

National Renewable Energy Laboratory

Suggested Citation

Reda, Ibrahim, Afshin Andreas, Peter Gotseff, Mark Kutchenreiter, and Martina Stoddard. 2019. *NREL Comparison of Absolute Cavity Pyrometers and Pyrometers Traceable to the World Infrared Standard Group and the Infrared Integrating Sphere: September 23-October 4, 2019*. Golden, CO: National Renewable Energy Laboratory. NREL/TP-1900-75119. <https://www.nrel.gov/docs/fy20osti/75119.pdf>.

**NREL is a national laboratory of the U.S. Department of Energy
Office of Energy Efficiency & Renewable Energy
Operated by the Alliance for Sustainable Energy, LLC**

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at www.nrel.gov/publications.

Contract No. DE-AC36-08GO28308

Technical Report
NREL/TP-1900-75119
October 2019

National Renewable Energy Laboratory
15013 Denver West Parkway
Golden, CO 80401
303-275-3000 • www.nrel.gov

NOTICE

This work was authored by the National Renewable Energy Laboratory, operated by Alliance for Sustainable Energy, LLC, for the U.S. Department of Energy (DOE) under Contract No. DE-AC36-08GO28308. Funding provided by U.S. Department of Energy Office of Energy Efficiency and Renewable Energy Solar Energy Technologies Office. The views expressed herein do not necessarily represent the views of the DOE or the U.S. Government.

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at www.nrel.gov/publications.

U.S. Department of Energy (DOE) reports produced after 1991 and a growing number of pre-1991 documents are available free via www.OSTI.gov.

Cover Photos by Dennis Schroeder: (clockwise, left to right) NREL 51934, NREL 45897, NREL 42160, NREL 45891, NREL 48097, NREL 46526.

NREL prints on paper that contains recycled content.

Acknowledgments

We sincerely appreciate the support of Solar Radiance Research Laboratory (SRRL) staff and National Renewable Energy Laboratory (NREL) management, the U.S. Department of Energy (DOE) Office of Energy Efficiency and Renewable Energy/Solar Energy Technologies Program, Environmental Research/Atmospheric Radiation Measurement Program, and NREL's Environment, Safety, Health, & Quality center (ESH&Q).

Table of Contents

1	Introduction.....	1
2	Instrument List.....	1
3	Measurement Equations	1
4	Results.....	2

List of Figures

Figure 1. ACPs' irradiance	3
Figure 2. Average ACPs' irradiance minus the irradiance measured by each ACP	4
Figure 3. Average ACPs' irradiance vs irradiance measured by PIR and KZ	5
Figure 4. Average of ACPs' irradiance minus irradiance measured by PIR and KZ.....	6
Figure 5. Water vapor content.....	7

List of Tables

Table 1. Average ACPs' Irradiance Minus the Irradiance Measured by Each ACP.....	7
--	---

1 Introduction

The comparison of the absolute cavity pyrgeometers (ACPs) with a precision infrared radiometer (PIR) pyrgeometer traceable to the World Infrared Standard Group (WISG) and a Kipp & Zonen (KZ) pyrgeometer traceable to WISG and the InfraRed Integrating Sphere Radiometer (IRIS) was held during the National Renewable Energy Laboratory (NREL) Pyrheliometer Comparison (NPC-2019) from September 23 to October 4, 2019. Data from all instruments was collected during nighttime clear sky conditions only. The irradiance measured by the ACPs, PIR, and CG4 pyrgeometers collected 30 seconds during the measurement runs. The measurement runs lasted for two hours, while the calibration runs lasted for 6 minutes. During the comparison, the average (av) irradiance difference measured by three ACPs varied from 0.2 W/m² to 0.8 W/m², with a standard deviation (sd) from 0.4 W/m² to 0.8 W/m². The average irradiance difference measured by the three ACPs minus the irradiance measured by the PIR was 7.1 W/m² and sd 1.0 W/m², and the average irradiance difference measured by the KZ traceable to WISG was 8.1 W/m² and sd 0.9 W/m², and average irradiance difference measured by the KZ traceable to IRIS was 4.0 W/m² and sd 0.9 W/m².

2 Instrument List

- Absolute Cavity Pyrgeometer (ACP): 57F3, 95F3, and 96F3.
- PIR pyrgeometer: 31197F3
- KZ pyrgeometer: CGR4 110390

3 Measurement Equations

ACP

$$W = \frac{K_1 * V_{tp} + (2 - \epsilon) * K_2 * W_r - (1 + \epsilon) * W_c}{\tau}$$

Where,

- W is the atmospheric longwave irradiance (W.m⁻²).
- K₁ is the reciprocal of the ACP's responsivity (W.m⁻².uV⁻¹).
- V_{tp} is the thermopile output voltage (uV).
- ε is the gold emittance.
- K₂ is the emittance of the black receiver surface.
- W_r is the receiver irradiance (W.m⁻²).
- W_c is the concentrator irradiance (W.m⁻²).
- τ is the ACP's throughput.

PIR

$$W = K_0 + K_1 * V_{tp} + K_2 * W_r + K_3 * (W_d - W_r)$$

Where,

- K_0 , K_1 , K_2 , and K_3 are the calibration coefficients.
- W_d is the dome irradiance, in W/m^2 .

KZ

$$W = \frac{V_{tp}}{C} + K_2 * W_c$$

Where C and K_2 are the calibration coefficients and W_c is the case irradiance.

4 Results

Figure 1 shows the irradiance of ACPs 57F3, 95F3, and 96F3, and Figure 2 shows the difference between the three ACPs' average irradiance and the irradiance measured by each ACP. Figure 3 shows the ACPs' average irradiance vs that of the PIR and KZ traceable to WISG and the KZ traceable to IRIS. Figure 4 shows the difference between the ACPs' average irradiance and the irradiance measured by the PIR and KZ with traceability to WISG and KZ with traceability to IRIS. Figure 5 is the water vapor content during the comparison. Table 1 shows that the difference between the ACPs average irradiance and ACP95F3, ACP57F3, and ACP96F3 varied from $-0.2 W/m^2$ to $0.8 W/m^2$, and uncertainty from $1 W/m^2$ to $1.7 W/m^2$ with 95% confidence level (U_{95}). As is shown in Table 1, the difference is $7.0 W/m^2$, $8.1 W/m^2$, and $4.0 W/m^2$ for the PIR and KZ with traceability to WISG and KZ with traceability to IRIS, and U_{95} of $7.3 W/m^2$, $8.3 W/m^2$, and $4.3 W/m^2$.

ACPs Irradiance from September 21 to 26, 2019




Figure 1. ACPs' irradiance

ACP Average Irradiance Minus Each ACP Irradiance
from September 21 to 26, 2019




Figure 2. Average ACPs' irradiance minus the irradiance measured by each ACP

ACPs Average Irradiance vs PIR and KZ Irradiance
from September 21 to 26, 2019




Figure 3. Average ACPs' irradiance vs irradiance measured by PIR and KZ




Figure 4. Average of ACPs' irradiance minus irradiance measured by PIR and KZ

PWV from September 21 to 26, 2019




Figure 5. Water vapor content

Table 1. Average ACPs' Irradiance Minus the Irradiance Measured by Each ACP

W/m ²	Wav-W95F3	Wav-W57F4	Wav-W96F5	Wav-W PIR	Wav-W(WISG)	Wav-W(IRIS)
av	-0.2	0.8	-0.5	7.0	8.1	4.0
sd	0.7	0.4	0.8	1.0	0.9	0.9
U ₉₅	1.5	1.0	1.7	7.3	8.3	4.3
nrdg	870	870	870	970	870	870