



# **NREL Comparison of Absolute Cavity Pyrometers, InfraRed Integrating Sphere, and Pyrometers Traceable to World Infrared Standard Group: September 26–October 7, 2022**

Ibrahim Reda, Afshin Andreas, Mark Kutchenreiter,  
Martina Stoddard, and Aron Habte

*National Renewable Energy Laboratory*

**NREL is a national laboratory of the U.S. Department of Energy  
Office of Energy Efficiency & Renewable Energy  
Operated by the Alliance for Sustainable Energy, LLC**

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at [www.nrel.gov/publications](http://www.nrel.gov/publications).

Contract No. DE-AC36-08GO28308

**Technical Report**  
NREL/TP-1900-84218  
November 2022



# **NREL Comparison of Absolute Cavity Pyrgometers, InfraRed Integrating Sphere, and Pyrgometers Traceable to World Infrared Standard Group: September 26–October 7, 2022**

Ibrahim Reda, Afshin Andreas, Mark Kutchenreiter,  
Martina Stoddard, and Aron Habte

*National Renewable Energy Laboratory*

## **Suggested Citation**

Reda, Ibrahim, Afshin Andreas, Mark Kutchenreiter, Martina Stoddard, Aron Habte. 2022. *NREL Comparison of Absolute Cavity Pyrgometers, InfraRed Integrating Sphere, and Pyrgometers Traceable to World Infrared Standard Group: September 26–October 7, 2022*. Golden, CO: National Renewable Energy Laboratory. NREL/TP-1900-84218. <https://www.nrel.gov/docs/fy23osti/84218.pdf>.

**NREL is a national laboratory of the U.S. Department of Energy  
Office of Energy Efficiency & Renewable Energy  
Operated by the Alliance for Sustainable Energy, LLC**

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at [www.nrel.gov/publications](http://www.nrel.gov/publications).

Contract No. DE-AC36-08GO28308

**Technical Report**  
NREL/TP-1900-84218  
November 2022

National Renewable Energy Laboratory  
15013 Denver West Parkway  
Golden, CO 80401  
303-275-3000 • [www.nrel.gov](http://www.nrel.gov)

## NOTICE

This work was authored by the National Renewable Energy Laboratory, operated by Alliance for Sustainable Energy, LLC, for the U.S. Department of Energy (DOE) under Contract No. DE-AC36-08GO28308. Funding provided by U.S. Department of Energy Office of Energy Efficiency and Renewable Energy Solar Energy Technologies Office. The views expressed herein do not necessarily represent the views of the DOE or the U.S. Government.

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at [www.nrel.gov/publications](http://www.nrel.gov/publications).

U.S. Department of Energy (DOE) reports produced after 1991 and a growing number of pre-1991 documents are available free via [www.OSTI.gov](http://www.OSTI.gov).

*Cover Photos by Dennis Schroeder: (clockwise, left to right) NREL 51934, NREL 45897, NREL 42160, NREL 45891, NREL 48097, NREL 46526.*

NREL prints on paper that contains recycled content.

## Acknowledgments

We sincerely appreciate the support of Solar Radiance Research Laboratory (SRRL) staff and National Renewable Energy Laboratory (NREL) management, the U.S. Department of Energy (DOE) Office of Energy Efficiency and Renewable Energy/Solar Energy Technologies Program, Environmental Research/Atmospheric Radiation Measurement Program, and NREL's Environment, Safety, Health, & Quality center (ESH&Q).

# Table of Contents

1	Introduction.....	1
2	Instrument List.....	1
3	Measurement Equations .....	1
4	Results.....	2

## List of Figures

Figure 1. ACPs and IRIS5 irradiance.....	3
Figure 2. ACP95F3 irradiance and irradiance measured by pyrgeometers.....	4
Figure 3. ACP95F3 thermopile output voltage.....	5
Figure 4. Water vapor content.....	6

## List of Tables

Table 1. ACP95F3 Irradiance Minus the Irradiance Measured by all radiometers.....	6
---	---

# 1 Introduction

The comparison of the absolute cavity pyrgeometers (ACPs) with the InfraRed Integrating Sphere (IRIS), Eppley Precision Infrared Radiometer (PIR) pyrgeometers, and Kipp & Zonen (KZ) pyrgeometers traceable to the World Infrared Standard Group (WISG) was held during NREL ACP and IRIS Comparisons (NAIC) from September 26 to October 7, 2022. Data from all instruments was collected during nighttime clear sky conditions only. The irradiance measured by the ACPs is collected in 30 seconds intervals during the measurement period of two hours, and 10 seconds intervals during the calibration period of 6 minutes.

During the comparison, the average (av) irradiance difference measured by ACPs and IRIS varied from -0.80 W/m<sup>2</sup> to 0.29 W/m<sup>2</sup> and standard deviation (sd) from 0.98 W/m<sup>2</sup> to 1.78 W/m<sup>2</sup>. The average irradiance difference measured by ACP95F3 minus the irradiance measured by all pyrgeometers varied from 2.07 to 5.03 W/m<sup>2</sup> with sd from 2.64 W/m<sup>2</sup> to 2.67 W/m<sup>2</sup>.

## Instrument List

- Absolute Cavity Pyrgeometer:
  - ACP20F3: DOE-Atmospheric atmospheric Radiation program (ARM)
  - ACP21F3: NREL
  - ACP57F3: German Weather Service (DWD)
  - ACP95F3: NREL
  - ACP96F3: Physikalisch-Meteorologisches Observatorium Davos—World Radiation Center (PMOD/WRC)
- InfraRed Integrating Sphere:
  - IRIS5: DWD
- PIR pyrgeometer: 31205F3, 31197F3, 36362F3 (NREL)
- KZ pyrgeometer CGR4: 060881(NREL), FT005(PMOD/WRC)

# 2 Measurement Equations

## ACP

$$W = \frac{K_1 * V_{tp} + (2 - \epsilon) * K_2 * W_r - (1 + \epsilon) * W_c}{\tau}$$

Where,

- W is the atmospheric longwave irradiance (W.m<sup>-2</sup>).
- K<sub>1</sub> is the reciprocal of the ACP's responsivity (W.m<sup>-2</sup>.uV<sup>-1</sup>).
- V<sub>tp</sub> is the thermopile output voltage (uV).
- ε is the gold emittance.
- K<sub>2</sub> is the emittance of the black receiver surface.
- W<sub>r</sub> is the receiver irradiance (W.m<sup>-2</sup>).
- W<sub>c</sub> is the concentrator irradiance (W.m<sup>-2</sup>).
- τ is the ACP's throughput.

## IRIS

$$W = \frac{U * \cos(\theta)}{C(1 + dt(T - 293.15))} + k\sigma T^4$$

Where,

- U is the signal (V).
- $\theta$  is the signal phase measure by the lock-in amplifier (°).
- C is the responsivity ( $VW^{-1}m^2$ ).
- T is the IRIS temperature (K).
- k is the emissivity correction factor.
- dt is the temperature coefficient of the pyroelectric detector ( $W.m^{-2}$ ).
- $\sigma$  is the Stefan-Boltzmann constant ( $W.m^{-2}$ ).

## PIR&KZ (NREL)

$$W = K_1 * V_{tp} + K_2 * W_r + K_3 * (W_d - W_r)$$

Where,

- $K_0, K_1, K_2$ , and  $K_3$  are the calibration coefficients.
- $W_d$  is the dome irradiance, in  $W/m^2$ .

## KZ (PMOD)

$$W = \frac{V_{tp}}{C} (1 + K_1 * \sigma T_c^3) + K_2 * W_c - K_3 * (W_d - W_c)$$

Where C,  $K_1$ , and  $K_2$  are the calibration coefficients,  $W_d$  and  $W_c$  are the dome and case irradiance.

## 3 Results

Figure 1 shows the irradiance of ACPs and IRIS5. Figure 2 shows ACP95F3 irradiance and irradiance measured by pyrgeometers. Figure 3 shows ACP95F3 thermopile output voltage. Figure 4 shows the water vapor content. Figure 5 is the water vapor content during the comparison. Table 1 shows that the difference between ACP95F3 and ACP20F3, ACP21F3, ACP57F3, ACP96F3 and IRIS5 varied from -0.2  $W/m^2$  to 0.8  $W/m^2$ , and uncertainty from 1.36  $W/m^2$  to 3.56  $W/m^2$  with 95% confidence level ( $U_{95}$ ). As is shown in Table 1,  $U_{95}$  of the difference between ACP95F3 and all pyrgeometers varied from 5.68  $W/m^2$  to 7.3  $W/m^2$ .



### ACPs and IRIS5 Irradiance from September 23 to 30, 2022




Figure 1. ACPs and IRIS5 irradiance

## ACP95F3 versus pyrometers Irradiance from September 23 to 30, 2022





Figure 2. ACP95F3 irradiance and irradiance measured by pyrometers

### ACP95F3 thermopile output voltage from September 23 to 30, 2022



**Figure 3. ACP95F3 thermopile output voltage\***

*\* ACP thermopile output voltage is a good indication of how clear the sky is.*




Figure 4. Water vapor content

Table 1. ACP95F3 Irradiance Minus the Irradiance Measured by all radiometers

W/m <sup>2</sup>	ACP95F3- ACP20F3	ACP95F3- ACP21F3	ACP95F3- ACP57F3	ACP95F3- ACP96F3	ACP95F3- IRIS5	ACP95F3- 06881	ACP95F3- FT005	ACP95F3- 31197F3	ACP95F3- 36362F3	ACP95F3- 31205F3
av	-0.80	-0.05	0.29	0.23	0.20	2.07	1.49	4.54	5.03	4.57
sd	0.98	0.68	1.42	0.88	1.78	2.64	2.12	2.64	2.67	2.63
U <sub>95</sub>	2.11	1.36	2.86	1.78	3.56	5.68	4.50	6.96	7.34	6.97
nrdg	2647	2421	2401	251	476	467	207	467	467	467