



NREL Comparison of Absolute Cavity Pyrometers, InfraRed Integrating Sphere, and Pyrometers Traceable to World Infrared Standard Group: September 25-October 6, 2023

Ibrahim Reda, Afshin Andreas, Martina Stoddard,
Aaron Kepple, Shawn Jaker, and Aron Habte

National Renewable Energy Laboratory

**NREL is a national laboratory of the U.S. Department of Energy
Office of Energy Efficiency & Renewable Energy
Operated by the Alliance for Sustainable Energy, LLC**

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at www.nrel.gov/publications.

Contract No. DE-AC36-08GO28308

Technical Report
NREL/TP-1900-88171
November 2023



NREL Comparison of Absolute Cavity Pyrometers, InfraRed Integrating Sphere, and Pyrometers Traceable to World Infrared Standard Group: September 25-October 6, 2023

Ibrahim Reda, Afshin Andreas, Martina Stoddard,
Aaron Kepple, Shawn Jaker, and Aron Habte

National Renewable Energy Laboratory

Suggested Citation

Reda, Ibrahim, Afshin Andreas, Martina Stoddard, Aaron Kepple, Shawn Jaker, and Aron Habte. 2023. *NREL Comparison of Absolute Cavity Pyrometers, InfraRed Integrating Sphere, and Pyrometers Traceable to World Infrared Standard Group: September 25-October 6, 2023*. Golden, CO: National Renewable Energy Laboratory. NREL/TP-1900-88171. <https://www.nrel.gov/docs/fy24osti/88171.pdf>.

**NREL is a national laboratory of the U.S. Department of Energy
Office of Energy Efficiency & Renewable Energy
Operated by the Alliance for Sustainable Energy, LLC**

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at www.nrel.gov/publications.

Contract No. DE-AC36-08GO28308

Technical Report
NREL/TP-1900-88171
November 2023

National Renewable Energy Laboratory
15013 Denver West Parkway
Golden, CO 80401
303-275-3000 • www.nrel.gov

NOTICE

This work was authored by the National Renewable Energy Laboratory, operated by Alliance for Sustainable Energy, LLC, for the U.S. Department of Energy (DOE) under Contract No. DE-AC36-08GO28308. Funding provided by NREL's Environment, Safety, Health, & Quality center (ESH&Q). The views expressed herein do not necessarily represent the views of the DOE or the U.S. Government.

This report is available at no cost from the National Renewable Energy Laboratory (NREL) at www.nrel.gov/publications.

U.S. Department of Energy (DOE) reports produced after 1991 and a growing number of pre-1991 documents are available free via www.OSTI.gov.

Cover Photos by Dennis Schroeder: (clockwise, left to right) NREL 51934, NREL 45897, NREL 42160, NREL 45891, NREL 48097, NREL 46526.

NREL prints on paper that contains recycled content.

Acknowledgments

We sincerely appreciate the support of Solar Radiance Research Laboratory (SRRL) staff and National Renewable Energy Laboratory (NREL) management, the U.S. Department of Energy (DOE) Office of Energy Efficiency and Renewable Energy/Solar Energy Technologies Program, Environmental Research/Atmospheric Radiation Measurement Program, and NREL's Environment, Safety, Health, & Quality center (ESH&Q).

Table of Contents

| | | |
|---|-----------------------------|---|
| 1 | Introduction..... | 1 |
| 2 | Instrument List..... | 1 |
| 3 | Measurement Equations | 1 |
| 4 | Results..... | 2 |

List of Figures

| | |
|---|---|
| Figure 1. ACPs and IRIS5 irradiance..... | 3 |
| Figure 2. ACP95F3 irradiance and irradiance measured by pyrgeometers..... | 4 |
| Figure 3. ACP95F3 thermopile output voltage..... | 5 |
| Figure 4. Water vapor content..... | 6 |

List of Tables

| | |
|---|---|
| Table 1. ACP95F3 Irradiance Minus the Irradiance Measured by all radiometers..... | 6 |
|---|---|

1 Introduction

The comparison of the absolute cavity pyrgeometers (ACPs) with the InfraRed Integrating Sphere (IRIS), Eppley Precision Infrared Radiometer (PIR) pyrgeometers, and Kipp & Zonen (KZ) pyrgeometers traceable to the World Infrared Standard Group (WISG) was held during NREL ACP and IRIS Comparisons (NAIC) from September 25 to October 6, 2023. Data from all instruments was collected during nighttime clear sky conditions only. The irradiance measured by the ACPs is collected in 30 seconds intervals during the measurement period of two hours, and 10 seconds intervals during the calibration period of 6 minutes.

During the comparison, the average (av) irradiance difference measured by ACPs and IRIS9 varied from -0.75 W/m^2 to 0.76 W/m^2 , standard deviation (sd) from 0.78 W/m^2 to 1.04 W/m^2 , and uncertainty U_{95} from 1.96 W/m^2 to 2.07 W/m^2 . The average irradiance difference measured by ACP95F3 minus the irradiance measured by all pyrgeometers varied from 1.64 to 3.96 W/m^2 , sd from 1.70 W/m^2 to 1.86 W/m^2 , and uncertainty U_{95} from 3.78 W/m^2 to 5.42 W/m^2 . Note that from September 25th at 18:31 to September 29th at 5:30 ACP96F3 irradiance is calculated using Bruce, et al 2023 method.

Instrument List

- Absolute Cavity Pyrgeometer:
 - ACP20F3: DOE-Atmospheric atmospheric Radiation program (ARM)
 - ACP95F3: NREL
 - ACP96F3: Physikalisch-Meteorologisches Observatorium Davos—World Radiation Center (PMOD/WRC)
 - ACP10F3: Japan Meteorological Agency (JMA)
- InfraRed Integrating Sphere:
 - IRIS9: PMOD
- PIR pyrgeometer: 31197F3 (NREL)
- KZ pyrgeometer CGR4: 060881(NREL), 010567 (JMA, Japan), FT005(PMOD/WRC)

2 Measurement Equations

ACP

$$W = \frac{K_1 * V_{tp} + (2 - \epsilon) * K_2 * W_r - (1 + \epsilon) * W_c}{\tau}$$

Where,

- W is the atmospheric longwave irradiance (W.m^{-2}).
- K_1 is the reciprocal of the ACP's responsivity ($\text{W.m}^{-2}.\text{uV}^{-1}$).
- V_{tp} is the thermopile output voltage (uV).
- ϵ is the gold emittance.
- K_2 is the emittance of the black receiver surface.
- W_r is the receiver irradiance (W.m^{-2}) = $\sigma * (T_{\text{case}} + 0.0007074 * V_{tp})^4$, where T_{case} is the pyrgeometer case temperature in Kelvin.

- W_c is the concentrator irradiance ($W \cdot m^{-2}$).
- τ is the ACP's throughput.

IRIS

$$W = \frac{U * \text{Cos}(\theta)}{C(1 + dt(T - 293.15))} + k\sigma T^4$$

Where,

- U is the signal (V).
- Θ is the signal phase measure by the lock-in amplifier ($^\circ$).
- C is the responsivity ($VW^{-1}m^2$).
- T is the IRIS temperature (K).
- k is the emissivity correction factor.
- dt is the temperature coefficient of the pyroelectric detector (K^{-1}).
- σ is the Stefan-Boltzmann constant ($W \cdot m^{-2}$).

PIR&KZ (NREL)

$$W = K_1 * V_{tp} + K_2 * W_r + K_3 * (W_d - W_r)$$

Where,

- K_1 , K_2 , and K_3 are the calibration coefficients.
- W_d is the dome irradiance, in W/m^2 .

KZ (PMOD)

$$W = \frac{V_{tp}}{C} (1 + K_1 * \sigma T_c^3) + K_2 * W_c - K_3 * (W_d - W_c)$$

Where C, K_1 , and K_2 are the calibration coefficients, W_d and W_c are the dome and case irradiance.

3 Results

Figure 1 shows the irradiance of ACPs and IRIS9. Figure 2 shows ACP95F3 irradiance and irradiance measured by pyrgeometers. Figure 3 shows ACP95F3 thermopile output voltage. Figure 4 shows the water vapor content. Figure 5 is the water vapor content during the comparison. Table 1 shows that U_{95} varied from $1.96 W/m^2$ to $2.07 W/m^2$ for all ACPs and IRIS9, and U_{95} varied from U_{95} from $3.78 W/m^2$ to $5.42 W/m^2$ for all pyrgeometers.

ACPs and IRIS5 Irradiance from September 17 to 29, 2023




Figure 1. ACPs and IRIS9 irradiance

ACP95F3 versus pyrgeometers Irradiance from September 17 to 29, 2023




Figure 2. ACP95F3 irradiance and irradiance measured by pyrgeometers

ACP95F3 thermopile output voltage from September 17 to 29, 2023




Figure 3. ACP95F3 thermopile output voltage*

** ACP thermopile output voltage is a good indication of how clear the sky is.*

PWV from September 17 to 29, 2023




Figure 4. Water vapor content

Table 1. ACP95F3 Irradiance Minus the Irradiance Measured by all radiometers

| W/m ² | ACP95F3-ACP20F3 | ACP95F3-ACP10F3 | ACP95F3-ACP96F3 | ACP95F3-IRIS9 | ACP95F3-31197F3 | ACP95F3-060881 | ACP95F3-010567 | ACP95F3-FT005 |
|------------------|-----------------|-----------------|-----------------|---------------|-----------------|----------------|----------------|---------------|
| av | -0.14 | 0.76 | -0.75 | 0.04 | 3.78 | 1.64 | 3.96 | 3.38 |
| sd | 0.78 | 1.00 | 0.84 | 1.04 | 1.74 | 1.70 | 1.86 | 1.76 |
| U ₉₅ | 1.56 | 2.13 | 1.84 | 2.07 | 5.13 | 3.78 | 5.42 | 4.88 |
| nrdg | 6675 | 2062 | 3766 | 1402 | 1841 | 1841 | 1841 | 1841 |